

Deutsche Akkreditierungsstelle GmbH

Annex to the Accreditation Certificate D-K-15074-01-00 according to DIN EN ISO/IEC 17025:2005

Period of validity: 16.09.2016 to 15.09.2021

Date of issue: 16.09.2016

Holder of certificate:

Mahr GmbH
Carl-Mahr-Straße 1, 37073 Göttingen

Head: Dipl.-Ing. (FH) Christoph Müller

Deputy: Dipl.-Ing. Karl-Josef Gödecke
Andreas Imke

Accredited as calibration laboratory since: 02.04.1990

Calibration in the fields:

Dimensional quantities

Length

- **Roughness**
- **Form error**
- **Contours**
- **Stylus instruments** ^{a)}
- **Length measuring devices** ^{a)}

^{a)} also on-site calibration

Abbreviations used: see last page

Permanent Laboratory

Measured quantity / Calibration item	Range	Measurement conditions / procedure	Best measurement capability ¹⁾	Remarks
Length Groove depth P_t on depth setting standards	0.2 µm to 0.8 µm (7.9 µin) (31.5 µin) 0.8 µm to 1.5 µm (31.5 µin) (59.1 µin) 1.5 µm to 3.5 µm (59.1 µin) (137.8 µin) 3.5 µm to 12 µm (137.8 µin) (472.5 µin)	DIN EN ISO 3274:1998 DIN EN ISO 4287:2010 DIN EN ISO 5436-1:2000	0.03 µm (1.2 µin) 0.04 µm (1.6 µin) 0.04 µm (1.6 µin) 0.05 µm (2.0 µin)	
Roughness on extra fine roughness standards R_a		DIN 4768:1990 DIN EN ISO 3274:1998 DIN EN ISO 4287:2010 DIN EN ISO 4288:1998 DIN EN ISO 16610-21:2013		
R_z	0.025 µm to 0.1 µm (1.0 µin) (3.9 µin)		0.08 · R_a	
R_{max}	0.15 µm to 0.8 µm (5.9 µin) (31.5 µin) 0.15 µm to 0.8 µm (5.9 µin) (31.5 µin)		0.09 · R_z 0.10 · R_{max}	
Roughness on roughness standards R_a	0.1 µm to 4 µm (3.9 µin) (157.5 µin)	DIN 4768:1990 DIN EN ISO 3274:1998 DIN EN ISO 4287:2010 DIN EN ISO 4288:1998	0.05 · R_a	
R_z	0.8 µm to 20 µm (31.5 µin) (787.5 µin)	DIN EN ISO 16610-21:2013	0.05 · R_z	
R_{max}	0.8 µm to 20 µm (31.5 µin) (787.5 µin)		0.05 · R_{max}	
Roughness on roughness standards R_{pk} R_k R_{vk}	On surfaces in the range 0.1 µm ≤ R_a ≤ 4 µm (3.9 µin) (157.5 µin)	DIN 4768:1990 DIN EN ISO 13565-1:1998 DIN EN ISO 13565-2:1998	0.04 · R_z 0.05 · R_z 0.04 · R_z	Relative measuring uncertainty relative to R_z
M_{r1} M_{r2}	0.8 µm ≤ R_z ≤ 20 µm (31.5 µin) (787.5 µin)		4 % 6 %	Relative measuring uncertainty relative to 100% material ratio
Roughness on roughness standards R_a	0.1 µm to 4 µm (3.9 µin) (157.5 µin)	DIN 4768:1990 DIN EN ISO 3274:1998 DIN EN ISO 4287:2010 DIN EN ISO 4288:1998	0.03 · R_a	If necessary, the filter cutoff wavelength λ_c can be used one level lower or higher than as per ISO 4288:1998
R_z	0.8 µm to 20 µm (31.5 µin) (787.5 µin)	DIN EN ISO 16610-21:2013	0.03 · R_z	
R_{max}	0.8 µm to 20 µm (31.5 µin) (787.5 µin)		0.03 · R_{max}	

¹⁾ The best measurement capabilities are stated according to DAkkS-DKD-3 (EA-4/02). These are expanded uncertainties of measurement with a coverage probability of 95% and have a coverage factor of $k = 2$ unless stated otherwise. Uncertainties without unit are relative uncertainties referring to the measurement value unless stated otherwise.

Annex to the accreditation certificate D-K-15074-01-00

Measured quantity / Calibration item	Range	Measurement conditions / procedure	Best measurement capability ¹⁾	Remarks
Stylus instruments to DIN EN ISO 3274:1998		DAkkS-DKD-R 4-2 part 2:2010 DIN EN ISO 12179:2000		U_{standard} is the measuring uncertainty of the standards used. Smaller measuring ranges for which standards are available can also be calibrated.
P_t	0.2 μm to 12 μm (7.9 μin) to (472.5 μin)		$U_{\text{standard}} + 0.01 \mu\text{m}$ ($U_{\text{standard}} + 0.4 \mu\text{in}$)	
R_a	0.1 μm to 4 μm (3.9 μin) to (157.5 μin)		$U_{\text{standard}} + 0.01 \cdot R_a$ ($U_{\text{standard}} + 0.4 \mu\text{in} \cdot R_a$)	
R_z	0.8 μm to 20 μm (31.5 μin) to (787.5 μin)		$U_{\text{standard}} + 0.01 \cdot R_z$ ($U_{\text{standard}} + 0.4 \mu\text{in} \cdot R_z$)	
R_{\max}	0.8 μm to 20 μm (31.5 μin) to (787.5 μin)		$U_{\text{standard}} + 0.01 \cdot R_{\max}$ ($U_{\text{standard}} + 0.4 \mu\text{in} \cdot R_{\max}$)	
Roundness standards		DIN ISO 1101:2014		Diameter: 3 mm to 100 mm (0.12 to 3.94 in)
Roundness deviation	to 0.1 μm (3.9 μin)		0.025 μm (1.0 μin)	
Magnification standards				Diameter: 3 mm to 100 mm (0.12 to 3.94 in)
Roundness deviation for cylinder with flat area (flick)	0.5 μm to 20 μm (19.7 μin) to (787.5 μin)		0.05 $\mu\text{m} + 2.5 \cdot 10^{-2} \cdot RON_t$ (2.0 $\mu\text{in} + 2.5 \cdot 10^{-2} \cdot RON_t$)	
Magnification standards		DIN ISO 1101:2014		Diameter: 50 mm to 150 mm (1.97 to 5.91 in)
Roundness deviation Multi-wave standard	to 20 μm (787.5 μin)		0.1 $\mu\text{m} + 2.5 \cdot 10^{-2} \cdot RON_t$ (3.9 $\mu\text{in} + 2.5 \cdot 10^{-2} \cdot RON_t$)	
Cylinder square		DIN ISO 1101:2014		Diameter: 3 mm to 100 mm (0.12 to 3.94 in)
Roundness deviation	to 20 μm (787.5 μin)		0.1 $\mu\text{m} + 2.5 \cdot 10^{-2} \cdot RON_t$ (3.9 $\mu\text{in} + 2.5 \cdot 10^{-2} \cdot RON_t$)	
Straightness deviation of the generatrices	to 20 μm (787.5 μin)		0.2 $\mu\text{m} + 2.0 \cdot 10^{-2} \cdot STR_t$ (7.9 $\mu\text{in} + 2.0 \cdot 10^{-2} \cdot STR_t$)	Length: 10 mm to 400 mm (0.39 to 15.75 in)
Parallelism deviation of the generatrices	to 20 μm (787.5 μin)		0.3 $\mu\text{m} + 1.5 \cdot 10^{-2} \cdot PAR_t$ (11.8 $\mu\text{in} + 1.5 \cdot 10^{-2} \cdot PAR_t$)	RON_t = roundness deviation STR_t = Straightness deviation PAR_t = Parallelism deviation
Cylindricity deviation	to 20 μm (787.5 μin)		0.4 $\mu\text{m} + 3.0 \cdot 10^{-2} \cdot CYL_t$ (15.8 $\mu\text{in} + 3.0 \cdot 10^{-2} \cdot CYL_t$)	CYL_t = Cylindricity deviation
Contour standards		Substitution measurement with reference contour standard Procedure according to DIN ISO/TS 15530-3:2008		
X length			0.6 μm (23.6 μin)	
Lateral distances	5 mm to 100 mm (0.20 in) to (3.94 in)		0.75 μm (29.5 μin)	
Z length			0.75 μm (29.5 μin)	
Vertical distances	to 10 mm (0.39 in)		0.01 $^\circ$	
Radii	2 mm to 12 mm (0.079 in) to (0.47 in)	MK03/05:2014 Calibration with traceable electronic linear reference gauge	0.22 μm (8.7 μin)	
Angles	40 $^\circ$ to 135 $^\circ$			
Dial gauge checkers	to 100 mm (3.94 in)			

¹⁾ The best measurement capabilities are stated according to DAkkS-DKD-3 (EA-4/02). These are expanded uncertainties of measurement with a coverage probability of 95% and have a coverage factor of $k = 2$ unless stated otherwise. Uncertainties without unit are relative uncertainties referring to the measurement value unless stated otherwise.

Measured quantity / Calibration item	Range	Measurement conditions / procedure	Best measurement capability ¹⁾	Remarks
Horizontal Length measuring machines	0 mm to 1000 mm (0 µin) (39.37 in)	VDI/VDE/DGQ 2618 part 17.1:2014	0.08 µm + 0.7 · 10 ⁻⁶ · l (3.1 µin + 0.7 · 10 ⁻⁶ · l)	l = measured length The measurement uncertainty of the length measurement uncertainty in mechanical probing of gauge blocks and is valid for horizontal length measuring machines of the Mahr GmbH
	> 1000 mm to 2000 mm (> 39.37 in) (78.74 in)		0.1 µm + 0.5 · 10 ⁻⁶ · l (3.9 µin + 0.5 · 10 ⁻⁶ · l)	

On-site calibration

Measured quantity / Calibration item	Range	Measurement conditions / procedure	Best measurement capability ¹⁾	Remarks
Length Dial gauge checkers	to 100 mm (3.94 in)	MK03/05:2014 Calibration with traceable electronic linear reference gauge	0.22 µm (8.7 µin)	
Stylus instruments to DIN EN ISO 3274:1998 <i>Pt</i>	0.2 µm to 12 µm (7.9 µin) (472.5 µin)	DAkkS-DKD-R 4-2 part 2:2010 DIN EN ISO 12179:2000	$U_{\text{standard}} + 0.01 \mu\text{m}$ ($U_{\text{standard}} + 0.4 \mu\text{in}$)	U_{standard} is the measuring uncertainty of the standards used. Smaller measuring ranges for which standards are available can also be calibrated.
<i>Ra</i>	0.1 µm to 4 µm (3.9 µin) (157.5 µin)		$U_{\text{standard}} + 0.01 \cdot Ra$ ($U_{\text{standard}} + 0.4 \mu\text{in} \cdot Ra$)	
<i>Rz</i>	0.8 µm to 20 µm (31.5 µin) (787.5 µin)		$U_{\text{standard}} + 0.01 \cdot Rz$ ($U_{\text{standard}} + 0.4 \mu\text{in} \cdot Rz$)	
<i>Rmax</i>	0.8 µm to 20 µm (31.5 µin) (787.5 µin)		$U_{\text{standard}} + 0.01 \cdot Rmax$ ($U_{\text{standard}} + 0.4 \mu\text{in} \cdot Rmax$)	
Horizontal Length measuring machines	0 mm to 1000 mm (0 µin) (39.37 in)	VDI/VDE/DGQ 2618 part 17.1:2014	0.08 µm + 0.7 · 10 ⁻⁶ · l (3.1 µin + 0.7 · 10 ⁻⁶ · l)	l = measured length The measurement uncertainty of the length measurement uncertainty in mechanical probing of gauge blocks and is valid for horizontal length measuring machines of the Mahr GmbH
	> 1000 mm to 2000 mm (> 39.37 in) (78.74 in)		0.1 µm + 0.5 · 10 ⁻⁶ · l (3.9 µin + 0.5 · 10 ⁻⁶ · l)	

Abbreviation used:

DAkkS-DKD-R	Guideline on Deutsche Akkreditierungsstelle GmbH
VDI/VDE/DGQ 2618	VDI-Inspection of measuring and test equipment
MK	Calibration instruction Mahr

¹⁾ The best measurement capabilities are stated according to DAkkS-DKD-3 (EA-4/02). These are expanded uncertainties of measurement with a coverage probability of 95% and have a coverage factor of $k = 2$ unless stated otherwise. Uncertainties without unit are relative uncertainties referring to the measurement value unless stated otherwise.