

Deutsche Akkreditierungsstelle GmbH

Annex to the Accreditation Certificate D-K-15089-01-00 according to DIN EN ISO/IEC 17025:2018

Valid from: 04.12.2020

Date of issue 07.01.2021

Holder of certificate:

Perschmann Calibration GmbH
Hauptstraße 46d, 38110 Braunschweig

Calibration in the fields:

Dimensional quantities

Length

- Gauge blocks
- Length measuring instruments
- Line scales, distances
- Length measuring devices ^{a)}
- Diameter
- Form error
- Flatness ^{a)}
- Straightness ^{a)}
- Thread

Coordinate measuring technology

- Coordinate measuring machines ^{b)}

Mechanical quantities

- Torque ^{c)}
- Pressure
- Weighing instruments ^{a)}
- Material testing machines (MTM)
 - Hardness (MTM)

Electrical quantities

DC and low frequency

- DC voltage
- AC voltage
- DC current
- AC current
- DC resistance

Time and frequency

- Frequency

Thermodynamic quantities

Temperature quantities

- Temperature indicators and simulators
- Resistance thermometers
- Radiation thermometers
- Temperature transmitters, data loggers
- Thermocouples
- Direct reading thermometers

Humidity quantities

- Devices for relative humidity

^{a)} also on-site calibration

^{b)} only on-site calibration

^{c)} also on-site calibration and calibration in the mobile laboratory

The management system requirements in DIN EN ISO/IEC 17025 are written in language relevant to operations of calibration laboratories and operate generally in accordance with the principles of DIN EN ISO 9001.

*The certificate together with its annex reflects the status at the time of the date of issue. The current status of the scope of accreditation can be found in the database of accredited bodies of Deutsche Akkreditierungsstelle GmbH.
<https://www.dakks.de/en/content/accredited-bodies-dakks>*

Abbreviations used: see last page

Page 1 of 13

This document is a translation. The definitive version is the original German annex to the accreditation certificate.

Annex to the accreditation certificate D-K-15089-01-00

Within the measurands/calibration items marked with with *, the calibration laboratory is permitted, without being required to inform and obtain prior approval from DAKKS, to use calibration standards or equivalent calibration procedures listed here with different issue dates.

The calibration laboratory maintains a current list of all calibration standards / equivalent calibration procedures within the flexible scope of accreditation.

Permanent Laboratory

Calibration and Measurement Capabilities (CMC)

Measurement quantity / Calibration item	Range	Measurement conditions / procedure	Expanded uncertainty of measurement ¹⁾	Remarks
Length Gauge blocks * made of steel according to DIN EN ISO 3650:1999	0.5 mm to 100 mm featuring the nominal values of the standards	DKD-R 4-3 part 3.1:2018 Measurement of the deviation of the central length l_c from the nominal value l_n by comparison measurement	For the central length: $0.05 \mu\text{m} + 0.5 \cdot 10^{-6} \cdot l$ For the deviations f_0 and f_u from the central length: $0.05 \mu\text{m}$	Measuring surface quality as stated in QMH resp. in the test specifications l = gauge block length
	> 100 mm to 150 mm featuring the nominal values of the standards	Measurement of the deviations f_0 and f_u from the central length by 5 points comparison For the smallest meas- urement uncertainties, the wringability and the wringing characteristics of both measuring surfaces must be checked using an appropriate optical flat	For the central length: $0.05 \mu\text{m} + 0.7 \cdot 10^{-6} \cdot l$ For the deviations f_0 and f_u from the central length: $0.07 \mu\text{m}$	
Gauge blocks * made of ceramics according to DIN EN ISO 3650:1999	0.5 mm to 100 mm featuring the nominal values of the standards	Measurement of the deviations f_0 and f_u from the central length by 5 points comparison For the smallest meas- urement uncertainties, the wringability and the wringing characteristics of both measuring surfaces must be checked using an appropriate optical flat	For the central length: $0.07 \mu\text{m} + 0.6 \cdot 10^{-6} \cdot l$ For the deviations f_0 and f_u from the central length: $0.05 \mu\text{m}$	
	> 100 mm to 150 mm featuring the nominal values of the standards		For the central length: $0.07 \mu\text{m} + 0.8 \cdot 10^{-6} \cdot l$ For the deviations f_0 and f_u from the central length: $0.07 \mu\text{m}$	

¹⁾ The expanded uncertainties according to EA-4/02 M:2013 are part of CMC and are the best measurement uncertainties within accreditation. They have a coverage probability of approximately 95 % and have a coverage factor of $k = 2$ unless stated otherwise. Uncertainties without unit are relative uncertainties referring to the measurement value unless stated otherwise.

Annex to the accreditation certificate D-K-15089-01-00

Permanent Laboratory

Calibration and Measurement Capabilities (CMC)

Measurement quantity / Calibration item	Range	Measurement conditions / procedure	Expanded uncertainty of measurement ¹⁾	Remarks
Gauge blocks * made of steel with special cross section (round or square), also with drilling in the middle	0.5 mm to 100 mm	DKD-R 4-3 part 3.1:2018 Measurement of the deviation of the central length l_c from the nominal value l_n by comparison measurement	For the central length: $0.1 \mu\text{m} + 0.5 \cdot 10^{-6} \cdot l$ For the deviations f_0 and f_u from the central length: $0.1 \mu\text{m}$	At square gauge blocks with drillings the mean size is substituted by ANSI-ASME B89.1.9M measured between hole and front side
Gauge blocks * made of tungsten carbide according to DIN EN ISO 3650:1999	0.5 mm to 100 mm featuring the nominal values of the standards	Measurement of the deviations f_0 and f_u from the central length by 5 points comparison For the smallest measurement uncertainties, the wringability and the wringing characteristics of both measuring surfaces must be checked using an appropriate optical flat	For the central length: $0.08 \mu\text{m} + 1.2 \cdot 10^{-6} \cdot l$ For the deviations f_0 and f_u from the central length: $0.05 \mu\text{m}$	
Gauge blocks * made of steel	> 150 mm to 1000 mm in the nominal dimensions, which differ of the standard with a max. of 50 mm	DKD-R 4-3 part 3.1:2018 Measurement of the deviation of the central length l_c from the nominal value l_n by comparison measurement	For the central length: $0.2 \mu\text{m} + 0.7 \cdot 10^{-6} \cdot l$	
Setting ring gauges * made of steel Diameter	2 mm to 200 mm	DKD-R 4-3 part 4.1:2018	$0.4 \mu\text{m} + 2 \cdot 10^{-6} \cdot d$	The measurement uncertainty applies to the complete calibration of diameter, roundness, straightness and parallelism. For the calibration of the diameter without form measurement, the best measurement uncertainty increases by $0.2 \mu\text{m}$ d = measured diameter
Setting plug gauges * made of steel Diameter	1 mm to 200 mm	DKD-R 4-3 part 4.1:2018	$0.4 \mu\text{m} + 2 \cdot 10^{-6} \cdot d$	
Measuring pins * made of steel Diameter	0.17 mm to 20 mm	DKD-R 4-3 part 4.2:2018	$0.4 \mu\text{m}$	
Roundness deviation * of abovementioned rings, inside cylinders, plugs or outside cylinders	to 40 μm	DKD-R 4-3 part 4.1:2018	$0.2 \mu\text{m} + 1 \cdot 10^{-2} \cdot R_{ONt}$	Diameter: 2 mm to 200 mm
Straightness deviation * of abovementioned rings, inside cylinders, plugs or outside cylinders	to 10 μm	DKD-R 4-3 part 4.1:2018	$0.5 \mu\text{m}$	axial length: to 30 mm
Setting dimension *	25 mm to 900 mm	VDI/VDE/DGQ 2618 part 4.4:2009	$0.7 \mu\text{m} + 1,5 \cdot 10^{-6} \cdot l$	l = measured length
Caliper gauge *	5 mm to 170 mm	DKD-R 4-3 part 4.7:2018	$1.5 \mu\text{m} + 10 \cdot 10^{-6} \cdot l$	

¹⁾ The expanded uncertainties according to EA-4/02 M:2013 are part of CMC and are the best measurement uncertainties within accreditation. They have a coverage probability of approximately 95 % and have a coverage factor of $k = 2$ unless stated otherwise. Uncertainties without unit are relative uncertainties referring to the measurement value unless stated otherwise.

Annex to the accreditation certificate D-K-15089-01-00

Permanent Laboratory

Calibration and Measurement Capabilities (CMC)

Measurement quantity / Calibration item	Range	Measurement conditions / procedure	Expanded uncertainty of measurement ¹⁾	Remarks
Reference- and Setting gauge	to 12 mm	Annex F/43:2017-11	3 µm	Feeler gauge, gap gauge of plastic, delrin, teflon, brass or steal
Radius gauge	to 40 mm	Annex F/42:2018-01	3 µm	to 40 mm radii
Angle meter *	0° to 360°	DKD-R 4-3 part 7.2:2018	1' 30"	
Graduator	0° to 180°	Annex F/46:2017-11	12'	
Measuring tape Circumference tape measure	0 m to 50 m	Annex F/47-1:2017-12 Annex F/47-2:2017-12	$50 \mu\text{m} + 15 \cdot 10^{-6} \cdot l$	$l =$ measured length
Diameter tape measure	0 m to 10 m	Annex F/47-2:2017-12	$50 \mu\text{m} + 15 \cdot 10^{-6} \cdot d$	$d =$ measured diameter
Rules	0 m to 10 m	Annex F/47-3:2017-12 Annex F/47-4:2017-11	$50 \mu\text{m} + 15 \cdot 10^{-6} \cdot l$	$l =$ measured length Graduated metal rules, reference- and plotting scale, rules, folding rules
Calipers for external, internal and depth dimensions *	0 mm to 1000 mm	DKD-R 4-3 part 9.1:2018	$30 \mu\text{m} + 30 \cdot 10^{-6} \cdot l$	$l =$ measured length
Depth calipers *	0 mm to 1000 mm	DKD-R 4-3 part 9.2:2018	$30 \mu\text{m} + 30 \cdot 10^{-6} \cdot l$	
Height gauge *	0 mm to 1000 mm	DKD-R 4-3 part 9.3:2018	$20 \mu\text{m} + 10 \cdot 10^{-6} \cdot l$	with contact help
Micrometers *	0 mm to 600 mm	DKD-R 4-3 part 10.1:2018	$3 \mu\text{m} + 10 \cdot 10^{-6} \cdot l$	600 mm = final value of the measuring range
Indicating caliper micrometers *	0 mm to 100 mm	DKD-R 4-3 part 10.3:2018	$2 \mu\text{m} + 10 \cdot 10^{-6} \cdot l$	$l =$ measured length
Indicating caliper gap gauge	0 mm to 100 mm	Annex F/39:2017-12	$2 \mu\text{m} + 10 \cdot 10^{-6} \cdot l$	
Depth micrometers *	0 mm to 300 mm	Annex F/36:2018-05	$3 \mu\text{m} + 10 \cdot 10^{-6} \cdot l$	with interchangeable contact points
Internal micrometers with two-point contact *	25 mm to 950 mm	DKD-R 4-3 part 10.7:2018	$3.5 \mu\text{m} + 5 \cdot 10^{-6} \cdot d$	$d =$ measured diameter
Internal micrometers with jaws	5 mm to 100 mm	Annex F/37:2017-11	$5 \mu\text{m} + 5 \cdot 10^{-6} \cdot d$	
Internal micrometers with three-point contact *	3 mm to 200 mm	DKD-R 4-3 part 10.8:2018	$3 \mu\text{m} + 5 \cdot 10^{-6} \cdot d$	
Dial gauges with scales * Scale interval > 1 µm	to 100 mm	DKD-R 4-3 part 11.1:2018	$3 \mu\text{m} + 10 \cdot 10^{-6} \cdot l$	$l =$ measured length
Dial gauges with scales * Scale interval 1 µm	to 5 mm		1.5 µm	error of measurement y_i
			2 µm	deviation span f_e, f_{ges}, f_u, f_t and f_w

¹⁾ The expanded uncertainties according to EA-4/02 M:2013 are part of CMC and are the best measurement uncertainties within accreditation. They have a coverage probability of approximately 95 % and have a coverage factor of $k = 2$ unless stated otherwise. Uncertainties without unit are relative uncertainties referring to the measurement value unless stated otherwise.

Annex to the accreditation certificate D-K-15089-01-00

Permanent Laboratory

Calibration and Measurement Capabilities (CMC)

Measurement quantity / Calibration item	Range	Measurement conditions / procedure	Expanded uncertainty of measurement ¹⁾	Remarks
Dial gauges with scales * Scale interval > 1 µm	to 100 mm	VDI/VDE/DGQ 2618 part 11.1:2014	$3 \mu\text{m} + 10 \cdot 10^{-6} \cdot l$	l = measured length
Dial gauges with scales * Scale interval 1 µm	to 5 mm		1.5 µm	error of measurement y_i
			2 µm	deviation span $MPE_r, MPE_e, MPE_{ges},$ $MPE_{1/1}, MPE_{1/2}, MPE_{1/10},$ MPE_u
Dial gauges with digital display Numerical interval 0.1 µm	to 25 mm	Annex F/04-2:2014-12	$0.6 \mu\text{m} + 10 \cdot 10^{-6} \cdot l$	error of measurement y_i
			$0.8 \mu\text{m} + 10 \cdot 10^{-6} \cdot l$	deviation span f_e, f_i and f_w
Dial gauges with digital display Numerical interval 1 µm	to 100 mm		$1 \mu\text{m} + 10 \cdot 10^{-6} \cdot l$	error of measurement y_i
			$1.5 \mu\text{m} + 10 \cdot 10^{-6} \cdot l$	deviation span f_e, f_i and f_w
Dial indicators * Scale interval > 0.5 µm	to 3 mm	DKD-R 4-3 part 11.2:2018	0.6 µm	
Lever gauges *	to 1.6 mm	DKD-R 4-3 part 11.3:2018	1 µm	
Lever gauges for external measurements *	0 mm to 70 mm	DKD-R 4-3 part 12.1:2018	$7 \mu\text{m} + 10 \cdot 10^{-6} \cdot l$	up to a probe length of 200 mm
Lever gauges for internal measurements *	2.5 mm to 80 mm	DKD-R 4-3 part 13.1:2018	$7 \mu\text{m} + 10 \cdot 10^{-6} \cdot l$	l = measured length
Thickness gauges *	0 mm to 30 mm	DKD-R 4-3 part 12.1:2018	$7 \mu\text{m} + 10 \cdot 10^{-6} \cdot l$	l = measured length up to a measuring depth of 200 mm
Bore gauges with two-point contact *	1 mm to 3 mm	VDI/VDE/DGQ 2618 part 13.2:2005 (image 1)	0.8 µm	range of application: with gauge slider $d = 1.75 \text{ mm}$ to $d = 25 \text{ mm}$
		VDI/VDE/DGQ 2618 part 13.2:2005 (image 2)	0.8 µm	range of application: to $d = 300 \text{ mm}$
			1.2 µm	range of application: $d > 300 \text{ mm}$ to $d = 600 \text{ mm}$
		VDI/VDE/DGQ 2618 part 13.2:2005 (image 3)	0.8 µm	range of application: plug gauge to $d = 100 \text{ mm}$
Height gauges *	0 mm to 1000 mm	VDI/VDE/DGQ 2618 part 16.1:2009	$1.5 \mu\text{m} + 3 \cdot 10^{-6} \cdot l$	l = measured length
Deviation from straightness and Perpendicularity	to 30 µm	to 800 mm lead length	$2.5 \mu\text{m} + 1 \cdot 10^{-6} \cdot l_z$	l_z = lead length

¹⁾ The expanded uncertainties according to EA-4/02 M:2013 are part of CMC and are the best measurement uncertainties within accreditation. They have a coverage probability of approximately 95 % and have a coverage factor of $k = 2$ unless stated otherwise. Uncertainties without unit are relative uncertainties referring to the measurement value unless stated otherwise.

Annex to the accreditation certificate D-K-15089-01-00

Permanent Laboratory

Calibration and Measurement Capabilities (CMC)

Measurement quantity / Calibration item	Range	Measurement conditions / procedure	Expanded uncertainty of measurement ¹⁾	Remarks	
90° Squares Perpendicularity	to 30 µm	Annex F/12:2017-02 to 750 mm leg length	$2 \mu\text{m} + 2 \cdot 10^{-6} \cdot l_z$	l_z = length of form respectively position embodiment	
Deviation from straightness and parallelism	to 30 µm	to 750 mm length	$1 \mu\text{m} + 2 \cdot 10^{-6} \cdot l_z$		
Flatness deviation	to 30 µm	to 750 mm edge length	$1 \mu\text{m} + 2 \cdot 10^{-6} \cdot l_z$		
Deviation from flatness Horizontal flatness standard, e.g. surface plates as per DIN 876:1984	to 50 µm	Annex F/13:2020-10 to 2 m edge length electronic inclination measuring	$0.9 \mu\text{m} + 1.7 \cdot 10^{-6} \cdot l$	l = longest edge length of the measuring standard For calibration in the permanent calibration laboratory, the uncertainty increases starting by an edge length $l > 1$ m by a factor of 1.2	
Deviation from straightness Horizontal flatness standard, e.g. surface plates as per DIN 876:1984	to 50 µm	Annex F/13:2020-10 to 3 m edge length electronic inclination measuring	$2.2 \mu\text{m} + 1.7 \cdot 10^{-6} \cdot l$		
Thread gauges * (single-start and multi-start cylindrical and conical ex- ternal and internal threads with straight flanks, sym- metrical and asymmetrical profile)					
External thread	3 mm to 150 mm	Scanning method DKD-R 4-3 part 4.8:2018, Option 1 to option 4 (Specifying the thread angle α)		d = measured diameter	
Simple pitch diameter	Nominal diameter		$2.5 \mu\text{m} + 5 \cdot 10^{-6} \cdot d$		
Outside diameter			2 µm		
Core diameter or recess diameter			5 µm		
Lead or pitch	0.5 mm to 8 mm		1 µm		
Thread angle α	$\geq 27^\circ$		$(1.2 + 3 \text{ mm} / l_F)'$, but not lower at 6'		l_F = side length
Internal thread	3 mm to 160 mm				
Simple pitch diameter	Nominal diameter	Scanning method	$2.5 \mu\text{m} + 5 \cdot 10^{-6} \cdot d$		
Outside diameter or recess diameter		DKD-R 4-3 part 4.9:2018, Option 1 to option 4 (Specifying the thread angle α)	5 µm		
Core diameter			2 µm		
Lead or pitch	0.5 mm to 8 mm		1 µm		
Thread angle α	$\geq 27^\circ$		$(1.2 + 3 \text{ mm} / l_F)'$, but not lower at 6'		
Stand off	3 mm to 150 mm	Annex F/09-3:2020-08	50 µm		

¹⁾ The expanded uncertainties according to EA-4/02 M:2013 are part of CMC and are the best measurement uncertainties within accreditation. They have a coverage probability of approximately 95 % and have a coverage factor of $k = 2$ unless stated otherwise. Uncertainties without unit are relative uncertainties referring to the measurement value unless stated otherwise.

Annex to the accreditation certificate D-K-15089-01-00
Permanent Laboratory
Calibration and Measurement Capabilities (CMC)

Measurement quantity / Calibration item	Range	Measurement conditions / procedure	Expanded uncertainty of measurement ¹⁾	Remarks
Thread plug gauge * Simple pitch diameter	1.4 mm to 150 mm	VDI/VDE/DGQ 2618 part 4.8:2006, option 1 (Three wire procedure)	$2.5 \mu\text{m} + 7.5 \cdot 10^{-6} \cdot d$	d = measured diameter P_h = lead $P_h \geq 0.3 \text{ mm to } \leq 6 \text{ mm}$
Torque * Hand torque assembly tools	1 N·m to 1000 N·m	DIN EN ISO 6789-2:2017	$5 \cdot 10^{-3}$	only operated torque tools
Torque wrench calibration devices	4 N·m to 1000 N·m	DKD-R 3-8:2018	$2 \cdot 10^{-3}$	
Pressure * gauge pressure p_e	1 bar to 700 bar $\geq 700 \text{ bar to } 800 \text{ bar}$	DIN EN 837:1997 DKD-R 6-1:2014	0.2 bar 0.5 bar	Pressure medium: oil
gauge pressure p_e	1 bar to 30 bar		0.01 bar	Pressure medium: gas
Weighing instruments * Nonautomatic weighing instruments	to $\leq 50 \text{ kg}$	EURAMET Calibration Guide No. 18 version 4.0 (11/2015)	$1,2 \cdot 10^{-5}$	with weights OIML R 111-1:2004 according to the class F1
Material testing machines (MTM) Hardness (MTM) * Hardness Testers according to hardness scales Shore A, AO and D	0 Shore to 100 Shore	DIN ISO 7619-1:2012 DIN ISO 18898:2014	1 Shore	Direct measurement with reference standards for travel and power. Optical calibration of the geometrical measurements with optical and tactile coordinate measuring machines
Measuring range	to 2.5 mm		$6 \mu\text{m}$	
Diameter, radii, lengths	to 27 mm		$3.5 \mu\text{m}$	
Area	to 600 mm^2		$5 \mu\text{m}^2$	
Angle	28° to 37°		0.1°	
Elastic force	0 N to 44.5 N		0.5 % of final value	
Shore A, AO and D Measuring path standard	0.5 mm to 2.5 mm	Annex F/34:2020-04	$0.8 \mu\text{m}$	

¹⁾ The expanded uncertainties according to EA-4/02 M:2013 are part of CMC and are the best measurement uncertainties within accreditation. They have a coverage probability of approximately 95 % and have a coverage factor of $k = 2$ unless stated otherwise. Uncertainties without unit are relative uncertainties referring to the measurement value unless stated otherwise.

Annex to the accreditation certificate D-K-15089-01-00

Permanent Laboratory

Calibration and Measurement Capabilities (CMC)

Measurement quantity / Calibration item	Range	Measurement conditions / procedure	Expanded uncertainty of measurement ¹⁾	Remarks
Temperature quantities				
Temperature indicators for thermocouples *	-200 °C to 1300 °C	DKD-R 5-5:2018	0.5 K	Simulation of the thermo-electric voltage with multifunction generator (input in temperature units (°C)) Electric consideration of the reference junction
Temperature indicators for resistance thermometers with sensor type PT100 *	-100 °C to 800 °C		0.2 K	Simulation of the resistance value on multifunction calibrator (input in temperature units (°C))
Resistance thermometers and direct reading thermometers with resistance sensor *	-25 °C to 140 °C	DKD-R 5-1:2018 in temperature block calibrator	0.2 K	Comparison with resistance thermometers
	> 140 °C to 300 °C		0.4 K	
	> 300 °C to 400 °C		0.6 K	
	> 400 °C to 500 °C		0.8 K	
	0 °C	DKD-R 5-1:2018 Ice point	50 mK	
Radiation thermometers *	5 °C to 120 °C > 120 °C to 500 °C	Black body radiators VDI/VDE 3511 part 4.4:2005	1.5 K 3 K	Calibration with plate radiators
Thermometers for air temperature (Data loggers)	10 °C to 50 °C	Annex F/22-BS:2019-10 in climatic chambers	0.3 K	Comparison with resistance thermometers
Thermocouples *	-25 °C to 140 °C > 140 °C to 500 °C	DKD-R 5-3:2018 in temperature block calibrator	1 K 2.2 K	Comparison with resistance thermometers
Humidity quantities *				
Devices for relative humidity in air no psychrometer	7 % to 90 %	DKD-R 5-8:2019 in humidity generator temperature: 23 °C	1.5 %	Comparison with reference humidity sensor Measurement uncertainty given in percent relative humidity
DC and low frequency quantities				
DC voltage Measuring instruments	0 mV to 220 mV > 0.22 V to 2.2 V > 2.2 V to 11 V > 11 V to 22 V > 22 V to 220 V > 220 V to 1100 V	Annex F/23-1:2020-05	0.65 μV + 6.6 · 10 ⁻⁶ · U 1.2 μV + 4.7 · 10 ⁻⁶ · U 2.6 μV + 3.5 · 10 ⁻⁶ · U 4 μV + 3.5 · 10 ⁻⁶ · U 5 μV + 5 · 10 ⁻⁶ · U 0.4 mV + 6.5 · 10 ⁻⁶ · U	U = measuring value

¹⁾ The expanded uncertainties according to EA-4/02 M:2013 are part of CMC and are the best measurement uncertainties within accreditation. They have a coverage probability of approximately 95 % and have a coverage factor of $k = 2$ unless stated otherwise. Uncertainties without unit are relative uncertainties referring to the measurement value unless stated otherwise.

Annex to the accreditation certificate D-K-15089-01-00

Permanent Laboratory

Calibration and Measurement Capabilities (CMC)

Measurement quantity / Calibration item	Range	Measurement conditions / procedure	Expanded uncertainty of measurement ¹⁾	Remarks
DC voltage Sources	0 V to 0.2 V > 0.2 V to 2 V > 2 V to 20 V > 20 V to 200 V > 200 V to 1000 V	Annex F/29-1:2020-05	0.53 $\mu\text{V} + 2.9 \cdot 10^{-6} \cdot U$ 1 $\mu\text{V} + 2.7 \cdot 10^{-6} \cdot U$ 4.1 $\mu\text{V} + 3.0 \cdot 10^{-6} \cdot U$ 40 $\mu\text{V} + 4.5 \cdot 10^{-6} \cdot U$ 0.5 mV + 4.5 $\cdot 10^{-6} \cdot U$	U = measuring value
DC current Measuring instruments	0 μA to 220 μA > 0.22 mA to 2.2 mA > 2.2 mA to 22 mA > 22 mA to 220 mA > 0.22 A to 2.2 A > 2.2 A to 3 A > 3 mA to 11 A > 11 A to 20.5 A	Annex F/23-2:2020-05	6 nA + 40 $\cdot 10^{-6} \cdot I$ 9.3 nA + 34 $\cdot 10^{-6} \cdot I$ 40 nA + 35 $\cdot 10^{-6} \cdot I$ 0.7 $\mu\text{A} + 45 \cdot 10^{-6} \cdot I$ 13 $\mu\text{A} + 79 \cdot 10^{-6} \cdot I$ 31 $\mu\text{A} + 0.29 \cdot 10^{-3} \cdot I$ 0.39 mA + 0.39 $\cdot 10^{-3} \cdot I$ 0.58 mA + 0.78 $\cdot 10^{-3} \cdot I$	I = measuring value
DC current sources	10 μA to 200 μA > 0.2 mA to 2 mA > 2 mA to 20 mA > 20 mA to 200 mA > 0.2 A to 2 A > 2 A to 20 A	Annex F/29-2:2020-05	0.4 nA + 12 $\cdot 10^{-6} \cdot I$ 8.9 nA + 10 $\cdot 10^{-6} \cdot I$ 41 nA + 13 $\cdot 10^{-6} \cdot I$ 0.8 $\mu\text{A} + 36 \cdot 10^{-6} \cdot I$ 17 $\mu\text{A} + 0.17 \cdot 10^{-3} \cdot I$ 0.4 mA + 0.38 $\cdot 10^{-3} \cdot I$	I = measuring value
DC current clamps	0.2 A to < 10 A 10 A to 100 A > 100 A to 1000 A	Annex F/23-2:2020-05	10 mA + 2 $\cdot 10^{-3} \cdot I$ 0.1 A + 2 $\cdot 10^{-3} \cdot I$ 0.8 A + 2.5 $\cdot 10^{-3} \cdot I$	I = measuring value with current coil with 2, 10 and 50 windings
AC voltage Measuring instruments	0.02 V to 0.22 V > 0.22 V to 2.2 V > 2.2 V to 22 V > 22 V to 220 V > 220 V to 1100 V	Annex F/23-3:2020-05 40 Hz to 20 kHz 40 Hz to 20 kHz 40 Hz to 20 kHz 50 Hz to 1 kHz	4 $\mu\text{V} + 80 \cdot 10^{-6} \cdot U$ 8 $\mu\text{V} + 42 \cdot 10^{-6} \cdot U$ 50 $\mu\text{V} + 0.24 \cdot 10^{-3} \cdot U$ 0.64 mV + 60 $\cdot 10^{-6} \cdot U$ 3.5 mV + 70 $\cdot 10^{-6} \cdot U$	U = measuring value
AC voltage Sources	10 mV to 200 mV > 0.2 V to 2.0 V > 2.0 V to 20 V	Annex F/29-3:2020-05 40 Hz to 100 Hz > 100 Hz to 2 kHz > 2 kHz to 10 kHz > 10 kHz to 30 kHz 40 Hz to 100 Hz > 100 Hz to 2 kHz > 2 kHz to 10 kHz > 10 kHz to 30 kHz 40 Hz to 100 Hz > 100 Hz to 2 kHz > 2 kHz to 10 kHz > 10 kHz to 30 kHz	4 $\mu\text{V} + 0.11 \cdot 10^{-3} \cdot U$ 2 $\mu\text{V} + 0.11 \cdot 10^{-3} \cdot U$ 4 $\mu\text{V} + 0.13 \cdot 10^{-3} \cdot U$ 8 $\mu\text{V} + 0.34 \cdot 10^{-3} \cdot U$ 20 $\mu\text{V} + 90 \cdot 10^{-6} \cdot U$ 20 $\mu\text{V} + 75 \cdot 10^{-6} \cdot U$ 20 $\mu\text{V} + 0.11 \cdot 10^{-3} \cdot U$ 40 $\mu\text{V} + 0.22 \cdot 10^{-3} \cdot U$ 0.2 mV + 90 $\cdot 10^{-6} \cdot U$ 0.2 mV + 75 $\cdot 10^{-6} \cdot U$ 0.2 mV + 0.11 $\cdot 10^{-3} \cdot U$ 0. mV + 0.22 $\cdot 10^{-3} \cdot U$	

¹⁾ The expanded uncertainties according to EA-4/02 M:2013 are part of CMC and are the best measurement uncertainties within accreditation. They have a coverage probability of approximately 95 % and have a coverage factor of $k = 2$ unless stated otherwise. Uncertainties without unit are relative uncertainties referring to the measurement value unless stated otherwise.

Annex to the accreditation certificate D-K-15089-01-00

Permanent Laboratory

Calibration and Measurement Capabilities (CMC)

Measurement quantity / Calibration item	Range	Measurement conditions / procedure	Expanded uncertainty of measurement ¹⁾	Remarks
AC voltage Sources	> 20 V to 200 V	40 Hz to 100 Hz > 100 Hz to 2 kHz > 2 kHz to 10 kHz > 10 kHz to 30 kHz	$2 \text{ mV} + 90 \cdot 10^{-6} \cdot U$ $2 \text{ mV} + 75 \cdot 10^{-3} \cdot U$ $2 \text{ mV} + 0.11 \cdot 10^{-3} \cdot U$ $4 \text{ mV} + 0.22 \cdot 10^{-3} \cdot U$	U = measuring value
	> 200 V to 1000 V	40 Hz to 10 kHz > 10 kHz to 30 kHz	$20 \text{ mV} + 0.11 \cdot 10^{-3} \cdot U$ $40 \text{ mV} + 0.22 \cdot 10^{-3} \cdot U$	
AC current Measuring instruments	20 μ A to 220 μ A > 0.22 mA to 2.2 mA > 2.2 mA to 22 mA > 22 mA to 220 mA > 0.22 A to 2.2 A > 2.2 A to 3 A > 3 A to 11 A > 11 A to 20.5 A	Annex F/23-4:2020-05 40 Hz to 1 kHz 40 Hz to 1 kHz 40 Hz to 1 kHz 40 Hz to 1 kHz 40 Hz to 1 kHz 45 Hz to 1 kHz 45 Hz to 100 Hz 45 Hz to 100 Hz	$8 \text{ nA} + 0.12 \cdot 10^{-3} \cdot I$ $36 \text{ nA} + 0.1 \cdot 10^{-3} \cdot I$ $0.35 \text{ } \mu\text{A} + 0.1 \cdot 10^{-3} \cdot I$ $2.5 \text{ } \mu\text{A} + 0.1 \cdot 10^{-3} \cdot I$ $36 \text{ } \mu\text{A} + 0.24 \cdot 10^{-3} \cdot I$ $78 \text{ } \mu\text{A} + 0.47 \cdot 10^{-3} \cdot I$ $1.6 \text{ mA} + 0.78 \cdot 10^{-3} \cdot I$ $3.9 \text{ mA} + 1.2 \cdot 10^{-3} \cdot I$	I = measuring value
AC current clumps	0.2 A to < 10 A 10 A to 100 A > 100 A to 500 A > 500 A to 1000 A	Annex F/23-4:2020-05 50 Hz 50 Hz 50 Hz 50 Hz	$10 \text{ mA} + 2 \cdot 10^{-3} \cdot I$ $0.1 \text{ A} + 2 \cdot 10^{-3} \cdot I$ $0.4 \text{ A} + 2.5 \cdot 10^{-3} \cdot I$ $0.8 \text{ A} + 2.5 \cdot 10^{-3} \cdot I$	I = measuring value with current coil with 2, 10 and 50 windings
AC current sources	0.2 mA to 2.0 mA > 2.0 mA to 20 mA > 20 mA to 200 mA > 0.2 A to 2 A > 2 A to 20 A	Annex F/29-4:2020-05 10 Hz to 10 kHz 10 Hz to 10 kHz 10 Hz to 10 kHz 10 Hz to 2 kHz 10 Hz to 2 kHz	$0.2 \text{ } \mu\text{A} + 0.3 \cdot 10^{-3} \cdot I$ $2 \text{ } \mu\text{A} + 0.3 \cdot 10^{-3} \cdot I$ $20 \text{ } \mu\text{A} + 0.29 \cdot 10^{-3} \cdot I$ $0.2 \text{ mA} + 0.62 \cdot 10^{-3} \cdot I$ $2 \text{ mA} + 0.82 \cdot 10^{-3} \cdot I$	I = measuring value
DC resistance Measuring instruments 4-wire connection	0 Ω 1 Ω 10 Ω 100 Ω 1 k Ω 10 k Ω 100 k Ω	Annex F/23-5:2020-05	2.3 m Ω 2.4 m Ω 2.8 m Ω 4.8 m Ω 37 m Ω 0.37 Ω 3.7 Ω	Fixed resistors
DC resistances 4-wire connection	0 Ω to 2 Ω > 2 Ω to 20 Ω > 20 Ω to 200 Ω > 200 Ω to 2 k Ω > 2 k Ω to 20 k Ω	Annex F/30:2020-05	$0.11 \text{ m}\Omega + 3 \cdot 10^{-6} \cdot R$ $0.1 \text{ m}\Omega + 6.1 \cdot 10^{-6} \cdot R$ $91 \text{ } \mu\Omega + 7.8 \cdot 10^{-6} \cdot R$ $0.12 \text{ } \Omega + 0.59 \cdot 10^{-3} \cdot R$ $0.11 \text{ } \Omega + 4.4 \cdot 10^{-6} \cdot R$	R = measuring value
Measuring instruments 2-wire connection	1 M Ω 10 M Ω 100 M Ω	Annex F/23-5:2020-05	0.06 k Ω 1.8 k Ω 0.24 M Ω	Fixed resistors

¹⁾ The expanded uncertainties according to EA-4/02 M:2013 are part of CMC and are the best measurement uncertainties within accreditation. They have a coverage probability of approximately 95 % and have a coverage factor of $k = 2$ unless stated otherwise. Uncertainties without unit are relative uncertainties referring to the measurement value unless stated otherwise.

Annex to the accreditation certificate D-K-15089-01-00

Permanent Laboratory

Calibration and Measurement Capabilities (CMC)

Measurement quantity / Calibration item	Range	Measurement conditions / procedure	Expanded uncertainty of measurement ¹⁾	Remarks
DC resistance DC resistances 2-wire connection	> 20 kΩ to 200 kΩ > 200 kΩ to 2 MΩ > 2 MΩ to 20 MΩ > 20 MΩ to 200 MΩ > 200 MΩ to 2 GΩ	Annex F/30:2020-05	91 mΩ + 7.8 · 10 ⁻⁶ · R 0.12 kΩ + 0.77 · 10 ⁻³ · R 0.14 kΩ + 0.81 · 10 ⁻³ · R 10 kΩ + 0.12 · 10 ⁻³ · R 1 MΩ + 0.51 · 10 ⁻³ · R	R = measuring value
Time and frequency Frequency Measuring instruments	0.01 Hz to 120 Hz > 120 Hz to 1.2 kHz > 1.2 kHz to 12 kHz > 12 kHz to 120 kHz > 120 kHz to 1.2 MHz	Annex F/23-6:2020-05	12 mHz + 50 · 10 ⁻⁶ · F 0.12 Hz + 50 · 10 ⁻⁶ · F 1.2 Hz + 50 · 10 ⁻⁶ · F 12 Hz + 50 · 10 ⁻⁶ · F 0.12 kHz + 50 · 10 ⁻⁶ · F	F = measuring value
Revolution speed Revolution counter optical	120 min ⁻¹ to 100000 min ⁻¹	Annex F/24:2020-05	0.05 min ⁻¹ + 18 · 10 ⁻⁶ · n	direct optical excitation n = measuring value

On-site Calibration

Calibration and Measurement Capabilities (CMC)

Measurement quantity / Calibration item	Range	Measurement conditions / procedure	Expanded uncertainty of measurement ¹⁾	Remarks
Length Deviation from flatness Horizontal flatness standard, e.g. surface plates as per DIN 876:1984	to 50 μm	Annex F/13:2020-10 to 2 m edge length electronic inclination measuring	0.9 μm + 1.7 · 10 ⁻⁶ · l	l = longest edge length of the measuring standard
Deviation from straightness Horizontal flatness standard, e.g. surface plates as per DIN 876:1984	to 50 μm	Annex F/13:2020-10 to 3 m edge length electronic inclination measuring	2.2 μm + 1.7 · 10 ⁻⁶ · l	
Height gauges *	0 mm to 600 mm	VDI/VDE/DGQ 2618 part 16.1:2009	2.5 μm + 5 · 10 ⁻⁶ · l	l = measured length

¹⁾ The expanded uncertainties according to EA-4/02 M:2013 are part of CMC and are the best measurement uncertainties within accreditation. They have a coverage probability of approximately 95 % and have a coverage factor of $k = 2$ unless stated otherwise. Uncertainties without unit are relative uncertainties referring to the measurement value unless stated otherwise.

Annex to the accreditation certificate D-K-15089-01-00

On-site Calibration

Calibration and Measurement Capabilities (CMC)				
Measurement quantity / Calibration item	Range	Measurement conditions / procedure	Expanded uncertainty of measurement ¹⁾	Remarks
Coordinate measuring technology * Measuring projectors Measuring microscopes	Devices featuring a measuring surface with a face diagonal ≤ 530 mm	Calibration of metro-logical characteristics according to DKD-R 4-3 part 18.1:2018, and the following standards and guidelines DIN EN ISO 10360 VDI/VDE 2617		
		Determination of probing error <i>PS-ID(OT)</i> with a graduated scale made of glass according to VDI/VDE 2617 part 6.1:2007	0.4 μ m	Measuring projectors and measuring Microscopes with visual probing with crosshairs or electronic edge detection
		The error of indication for size measurement <i>E-ID(OT)</i> and <i>E-2D(OT)</i> is determined with a graduated scale made of glass according to VDI/VDE 2617 part 6.1:2007	0.5 μ m + 0.5 · 10 ⁻⁶ · L	L = measured length
Weighing instruments * Nonautomatic weighing instruments	to ≤ 50 kg	EURAMET Calibration Guide No. 18 version 4.0 (11/2015)	1 · 10 ⁻⁵	with weights OIML R 111-1:2004 according to the class F1
Torque * Hand torque assembly tools	1 N·m to 1000 N·m	DIN EN ISO 6789-2:2017	5 · 10 ⁻³	only operated torque tools

Mobile Laboratory

Calibration and Measurement Capabilities (CMC)				
Measurement quantity / Calibration item	Range	Measurement conditions / procedure	Expanded uncertainty of measurement ¹⁾	Remarks
Torque * Hand torque assembly tools	1 N·m to 1000 N·m	DIN EN ISO 6789-2:2017	5 · 10 ⁻³	only operated torque tools

¹⁾ The expanded uncertainties according to EA-4/02 M:2013 are part of CMC and are the best measurement uncertainties within accreditation. They have a coverage probability of approximately 95 % and have a coverage factor of $k = 2$ unless stated otherwise. Uncertainties without unit are relative uncertainties referring to the measurement value unless stated otherwise.

Abbreviations used:

CMC	Calibration and measurement capabilities (Kalibrier- und Messmöglichkeiten)
DIN	Deutsches Institut für Normung e.V.
DKD-R	Guideline of Deutschen Kalibrierdienstes (DKD), published by Physikalisch-Technischen Bundesanstalt
EURAMET	European Association of National Metrology Institutes
VDE	Verband der Elektrotechnik, Elektronik und Informationstechnik e.V.
VDI	Verein Deutscher Ingenieure e.V.
DGQ	Deutsche Gesellschaft für Qualität e.V.
Annex F	Calibration Guide of Perschmann Calibration GmbH

¹⁾ The expanded uncertainties according to EA-4/02 M:2013 are part of CMC and are the best measurement uncertainties within accreditation. They have a coverage probability of approximately 95 % and have a coverage factor of $k = 2$ unless stated otherwise. Uncertainties without unit are relative uncertainties referring to the measurement value unless stated otherwise.