

# Deutsche Akkreditierungsstelle GmbH

## Annex to the Accreditation Certificate D-K-15118-01-01 according to DIN EN ISO/IEC 17025:2018

Valid from: 22.10.2020

Date of issue 22.10.2020

Holder of certificate:

**Kessler QMP GmbH**  
**Nisterberger Weg 16, 57520 Friedewald**

Calibration in the fields:

### **Dimensional quantities**

#### **Length**

- Length gauges
- Diameter
- Form error
- Length measuring instruments <sup>b)</sup>
- Length measuring devices <sup>a)</sup>
- Flatness <sup>a)</sup>
- Straightness
- Line scales, distances
- Thread
- Gear quantities

### **Coordinate measuring technology**

- Application coordinate measuring machines
- Coordinate measuring machines <sup>c)</sup>

#### **Angle**

- Angle gauges
- Inclination measuring instruments

### **Mechanical quantities**

- Force
- Weighing instruments <sup>a)</sup>
- Torque <sup>a)</sup>
- Pressure <sup>a), b)</sup>

### **Thermodynamic quantities**

#### **Temperature quantities**

- Direct reading thermometers

<sup>a)</sup> also on-site-calibration

<sup>b)</sup> also calibration in the mobile laboratory

<sup>c)</sup> only on-site-calibration

*The management system requirements in DIN EN ISO/IEC 17025 are written in language relevant to operations of calibration laboratories and operate generally in accordance with the principles of DIN EN ISO 9001.*

*The certificate together with its annex reflects the status at the time of the date of issue. The current status of the scope of accreditation can be found in the database of accredited bodies of Deutsche Akkreditierungsstelle GmbH.  
<https://www.dakks.de/en/content/accredited-bodies-dakks>*

**Annex to the accreditation certificate D-K-15118-01-01**
**Permanent Laboratory**
**Calibration and Measurement Capabilities (CMC)**

Measurement quantity / Calibration item	Range	Measurement conditions / procedure	Expanded uncertainty of measurement <sup>1)</sup>	Remarks
<b>Länge</b> Gauge blocks made of steel according to DIN EN ISO 3650:1999  Gauge blocks made of ceramics according to DIN EN ISO 3650:1999  Gauge blocks made of tungsten carbide according to DIN EN ISO 3650:1999	0.5 mm to 100 mm  featuring the nominal values of the steel standards	VDI/VDE/DGQ 2618 part 3.1:2004 Measurement of the deviation of the central length $l_c$ from the nominal value $l_n$ by comparison measurement Measurement of the deviations $f_o$ and $f_u$ from the central length by 5 points comparison measurement For the smallest measurement uncertainties, the wringability and the wringing characteristics of both measuring surfaces must be checked using an appropriate optical flat	For the central length: $0.08 \mu\text{m} + 0.8 \cdot 10^{-6} \cdot l$  For the deviations $f_o$ and $f_u$ from the central length: $0.05 \mu\text{m}$	$l =$ gauge block length
			For the central length: $0.1 \mu\text{m} + 1.1 \cdot 10^{-6} \cdot l$  For the deviations $f_o$ and $f_u$ from the central length: $0.07 \mu\text{m}$	
			For the central length: $0.1 \mu\text{m} + 3.3 \cdot 10^{-6} \cdot l$  For the deviations $f_o$ and $f_u$ from the central length: $0.07 \mu\text{m}$	
Setting plug gauges Diameter	1 mm to 500 mm	VDI/VDE/DGQ 2618 part 4.1:2006, option 1, option 2  Option 3, option 4	$0.4 \mu\text{m} + 4 \cdot 10^{-6} \cdot d$  $0.8 \mu\text{m} + 3 \cdot 10^{-6} \cdot d$	$d =$ measured diameter
Roundness deviation	to 40 $\mu\text{m}$	VDI/VDE/DGQ 2618 part 4.1:2006	$0.3 \mu\text{m} + 2.5 \cdot 10^{-2} \cdot RONt$	
Straightness deviation	to 40 $\mu\text{m}$		$0.4 \mu\text{m} + 2.5 \cdot 10^{-2} \cdot STRt$	
Parallelism deviation	to 40 $\mu\text{m}$		$0.7 \mu\text{m} + 2.5 \cdot 10^{-2} \cdot PARt$	
Setting ring gauges Diameter	2 mm to 200 mm	VDI/VDE/DGQ 2618 part 4.1:2006 option 1, option 2  Option 3, option 4	$0.6 \mu\text{m} + 3 \cdot 10^{-6} \cdot d$  $0.8 \mu\text{m} + 3 \cdot 10^{-6} \cdot d$	$d =$ measured diameter
Roundness deviation	to 40 $\mu\text{m}$	VDI/VDE/DGQ 2618 part 4.1:2006	$0.3 \mu\text{m} + 2.5 \cdot 10^{-2} \cdot RONt$	
Straightness deviation	to 40 $\mu\text{m}$		$0.4 \mu\text{m} + 2.5 \cdot 10^{-2} \cdot STRt$	
Parallelism deviation	to 40 $\mu\text{m}$		$0.7 \mu\text{m} + 2.5 \cdot 10^{-2} \cdot PARt$	
Roundness deviation	to 40 $\mu\text{m}$	TK 40:2016-12	$0.3 \mu\text{m} + 2.5 \cdot 10^{-2} \cdot RONt$	
Straightness deviation	to 40 $\mu\text{m}$		$0.4 \mu\text{m} + 2.5 \cdot 10^{-2} \cdot STRt$	
Parallelism deviation	to 40 $\mu\text{m}$		$0.7 \mu\text{m} + 2.5 \cdot 10^{-2} \cdot PARt$	
Measuring pins / Pins for screw threads Diameterr	1 mm to 50 mm	VDI/VDE/DGQ 2618 part 4.2:2007, option 1, option 2	0.6 $\mu\text{m}$	
	0.17 mm to 50 mm	Option 3	0.8 $\mu\text{m}$	

<sup>1)</sup> The expanded uncertainties according to EA-4/02 M:2013 are part of CMC and are the best measurement uncertainties within accreditation. They have a coverage probability of approximately 95 % and have a coverage factor of  $k = 2$  unless stated otherwise. Uncertainties without unit are relative uncertainties referring to the measurement value unless stated otherwise.

Date of issue: 22.10.2020

Valid from: 22.10.2020

Page 2 of 17

Annex to the accreditation certificate D-K-15118-01-01

Permanent Laboratory

Calibration and Measurement Capabilities (CMC)

Measurement quantity / Calibration item	Range	Measurement conditions / procedure	Expanded uncertainty of measurement <sup>1)</sup>	Remarks
Roundness deviation	to 40 µm	VDI/VDE/DGQ 2618 part 4.1:2006	$0.3 \mu\text{m} + 2.5 \cdot 10^{-2} \cdot RON_t$	
Straightness deviation	to 40 µm		$0.4 \mu\text{m} + 2.5 \cdot 10^{-2} \cdot STR_t$	
Taper gauges Taper plug gauges and Taper ring gauges Diameter	1 mm to 200 mm	VDI/VDE/DGQ 2618 part 4.12:2007, option 1 and 2	$0.8 \mu\text{m} + 3 \cdot 10^{-6} \cdot d$	<i>d</i> = measured diameter
Angular deviation	to 5'		2''	
Roundness deviation	to 40 µm		$0.3 \mu\text{m} + 2.5 \cdot 10^{-2} \cdot RON_t$	
Straightness deviation	to 40 µm		0.6 µm	
Gap gauges	5 mm to 200 mm	VDI/VDE/DGQ 2618 part 4.7:2005	$3 \mu\text{m} + 10 \cdot 10^{-6} \cdot l$	<i>l</i> = measured length
Straight edges Flatness deviation	to 1000 mm	VDI/VDE/DGQ 2618 part 5.1:2013	$1.1 \mu\text{m} + 1.6 \cdot 10^{-6} \cdot l$	
Parallelism deviation			$2.2 \mu\text{m} + 3.2 \cdot 10^{-6} \cdot l$	
Knife straight edges	to 1000 mm	VDI/VDE/DGQ 2618 part 5.2:2013	$1.1 \mu\text{m} + 1.6 \cdot 10^{-6} \cdot l$	
Surface plates Flatness deviation	to 50 µm	VDI/VDE/DGQ 2618 part 6.2:2014 to 8 m edge length	$1.3 \mu\text{m} + 1.3 \cdot 10^{-6} \cdot l$	<i>l</i> = measured length with inclination measuring instruments
Steel squares Flatness deviation	to 1000 mm	VDI/VDE/DGQ/DKD 2618 part 7.1:2019	$1.1 \mu\text{m} + 0.8 \cdot 10^{-6} \cdot l$	<i>l</i> = measured length
Angular deviation			$3.1 \mu\text{m} + 5 \cdot 10^{-6} \cdot l$	
Centring angles Flatness deviation	to 1000 mm	TK 90:2020-01	$1.1 \mu\text{m} + 0.8 \cdot 10^{-6} \cdot l$	
Angular deviation			$3.1 \mu\text{m} + 5 \cdot 10^{-6} \cdot l$	
Protractors Flatness deviation	0° to 360°	VDI/VDE/DGQ 2618 part 7.2:2008	5 µm	
Parallelism deviation			5 µm	
Angle Scale interval 5'	0° to 360°		4'	
Scale interval 1°	0° to 180°		24'	
Calipers for external, internal and depth dimensions	0 mm to 300 mm	VDI/VDE/DGQ 2618 part 9.1:2006	$30 \mu\text{m} + 30 \cdot 10^{-6} \cdot l$	
	> 300 mm to 1500 mm		$50 \mu\text{m} + 30 \cdot 10^{-6} \cdot l$	
	> 1500 mm to 3000 mm		$70 \mu\text{m} + 30 \cdot 10^{-6} \cdot l$	
Depth calipers	0 mm to 300 mm	VDI/VDE/DGQ 2618 part 9.2:2006	$30 \mu\text{m} + 30 \cdot 10^{-6} \cdot l$	
	> 300 mm to 1500 mm		$50 \mu\text{m} + 30 \cdot 10^{-6} \cdot l$	

<sup>1)</sup> The expanded uncertainties according to EA-4/02 M:2013 are part of CMC and are the best measurement uncertainties within accreditation. They have a coverage probability of approximately 95 % and have a coverage factor of  $k = 2$  unless stated otherwise. Uncertainties without unit are relative uncertainties referring to the measurement value unless stated otherwise.

Date of issue: 22.10.2020

Valid from: 22.10.2020

Annex to the accreditation certificate D-K-15118-01-01

Permanent Laboratory

Calibration and Measurement Capabilities (CMC)

Measurement quantity / Calibration item	Range	Measurement conditions / procedure	Expanded uncertainty of measurement <sup>1)</sup>	Remarks
Height calipers with analogue display	0 mm to 600 mm	VDI/VDE/DGQ 2618 part 9.3:2006	$30 \mu\text{m} + 30 \cdot 10^{-6} \cdot l$	<i>l</i> = final value of the measuring range
			with digital display	
Micrometers	0 mm to 100 mm	VDI/VDE/DGQ 2618 part 10.1:2001	$3 \mu\text{m} + 10 \cdot 10^{-6} \cdot l$	
	> 100 mm to 500 mm		$4 \mu\text{m} + 10 \cdot 10^{-6} \cdot l$	
	> 500 mm to 1000 mm	TK 2:2020-02	$5 \mu\text{m} + 10 \cdot 10^{-6} \cdot l$	
	> 1000 mm to 1500 mm		$6 \mu\text{m} + 10 \cdot 10^{-6} \cdot l$	
Reference gauges for micrometers	25 mm to 500 mm	VDI/VDE/DGQ 2618 part 4.4:2009	$2 \mu\text{m} + 5 \cdot 10^{-6} \cdot l$	
	> 500 mm to 1500 mm		$3 \mu\text{m} + 5 \cdot 10^{-6} \cdot l$	
Micrometers with interchangeable inserts	0 mm to 100 mm	VDI/VDE/DGQ 2618 part 10.2:2010	$3 \mu\text{m} + 10 \cdot 10^{-6} \cdot l$	
	> 100 mm to 300 mm		$5 \mu\text{m} + 10 \cdot 10^{-6} \cdot l$	
Micrometers with dial indicator	0 mm to 100 mm	VDI/VDE/DGQ 2618 part 10.3:2002	$3 \mu\text{m} + 10 \cdot 10^{-6} \cdot l$	
Micrometers heads	0 mm to 100 mm	VDI/VDE/DGQ 2618 part 10.4:2008	$3 \mu\text{m} + 10 \cdot 10^{-6} \cdot l$	<i>l</i> = measured length
	> 100 mm to 500 mm		$4 \mu\text{m} + 10 \cdot 10^{-6} \cdot l$	
	> 500 mm to 1000 mm		$5 \mu\text{m} + 10 \cdot 10^{-6} \cdot l$	
Depth micrometers	0 mm to 100 mm	VDI/VDE/DGQ 2618 part 10.5:2010	$3 \mu\text{m} + 10 \cdot 10^{-6} \cdot l$	
	> 100 mm to 500 mm		$4 \mu\text{m} + 10 \cdot 10^{-6} \cdot l$	
	> 500 mm to 1000 mm		$5 \mu\text{m} + 10 \cdot 10^{-6} \cdot l$	
Internal micrometers with two-point contact	25 mm to 100 mm	VDI/VDE/DGQ 2618 part 10.7:2010	$3 \mu\text{m} + 10 \cdot 10^{-6} \cdot l$	
	> 100 mm to 500 mm		$4 \mu\text{m} + 10 \cdot 10^{-6} \cdot l$	
	> 500 mm to 1000 mm		$5 \mu\text{m} + 10 \cdot 10^{-6} \cdot l$	
	> 1000 mm to 1500 mm		$6 \mu\text{m} + 10 \cdot 10^{-6} \cdot l$	
Extensions for internal micrometers with two-point contact	25 mm to 500 mm	VDI/VDE/DGQ 2618 part 10.7:2010	$2 \mu\text{m} + 5 \cdot 10^{-6} \cdot l$	
	> 500 mm to 1500 mm		$3.5 \mu\text{m} + 5 \cdot 10^{-6} \cdot l$	
Internal micrometers with three-point contact	3 mm to 200 mm	VDI/VDE/DGQ 2618 part 10.8:2002	$4 \mu\text{m} + 10 \cdot 10^{-6} \cdot d$	<i>d</i> = measured diameter
Internal measuring instruments	3 mm to 200 mm	TK 57:2017-03	$2 \mu\text{m} + 10 \cdot 10^{-6} \cdot d$	
Dial gauges Scale interval > 1 $\mu\text{m}$	to 100 mm	VDI/VDE/DGQ 2618 part 11.1:2014	$3 \mu\text{m} + 10 \cdot 10^{-6} \cdot l$	<i>l</i> = measured length
			Scale interval $\leq 1 \mu\text{m}$	
Dial indicators	to 3 mm	VDI/VDE/DGQ 2618 part 11.2:2002	1.1 $\mu\text{m}$	

<sup>1)</sup> The expanded uncertainties according to EA-4/02 M:2013 are part of CMC and are the best measurement uncertainties within accreditation. They have a coverage probability of approximately 95 % and have a coverage factor of  $k = 2$  unless stated otherwise. Uncertainties without unit are relative uncertainties referring to the measurement value unless stated otherwise.

**Annex to the accreditation certificate D-K-15118-01-01**

**Permanent Laboratory**

**Calibration and Measurement Capabilities (CMC)**

Measurement quantity / Calibration item	Range	Measurement conditions / procedure	Expanded uncertainty of measurement <sup>1)</sup>	Remarks
Lever gauges	to 1.6 mm	VDI/VDE/DGQ 2618 part 11.3:2002	1.2 µm	
Lever gauges (quicktests) for external measurements	0 mm to 100 mm	VDI/VDE/DGQ 2618 part 12.1:2005	6 µm	
Thickness gauges Scale interval 1 µm	0 mm to 30 mm	VDI/VDE/DGQ 2618 part 12.1:2005	$1.3 \mu\text{m} + 10 \cdot 10^{-6} \cdot l$	l = measured length
Scale interval 10 µm			$6 \mu\text{m} + 10 \cdot 10^{-6} \cdot l$	
Lever gauges (quicktests) for internal measurements	2.5 mm to 500 mm	VDI/VDE/DGQ 2618 part 13.1:2005	6 µm	
Bore gauges with two-point contact Form I - III	1 mm to 800 mm	VDI/VDE/DGQ 2618 part 13.2:2005	$1.8 \mu\text{m} + 10 \cdot 10^{-6} \cdot l$	Measuring length up to 3 mm
Electical probe and measuring device	0 mm to 10 mm	VDI/VDE/DGQ 2618 part 14.1:2010	1.4 µm	
Feeler gauges	10 µm to 2 mm	DIN 2275:2014	1 µm	
Measuring tape, Circumference tape measure	0 m to 100 m	TK 85:2020-01	$56 \mu\text{m} + 46 \cdot 10^{-6} \cdot l$	l = measured length
Rules	0 m to 5 m	TK 85:2020-01	$56 \mu\text{m} + 46 \cdot 10^{-6} \cdot l$	Graduated metal rules, reference- and plotting scale, rules, folding rules
Diameter tape measure	0 m to 10 m	TK 85:2020-01	$56 \mu\text{m} + 46 \cdot 10^{-6} \cdot l$	
Height gauges	0 mm to 1000 mm	VDI/VDE/DGQ 2618 part 16.1:2009	$1.7 \mu\text{m} + 1.2 \cdot 10^{-6} \cdot l$	l = measured length
Setting dimension for height gauges	to 20 mm	TK 89:2020-01	0.5 µm	
Height gauges	0 mm to 1000 mm	VDI/VDE/DGQ 2618 part 16.1:2009	0.9 µm	till 1000 mm lead length
Deviation from straightness and perpendicularity	to 40 µm		4 µm	
Horizontal length measuring device	0 mm to 5000 mm	VDI/VDE/DGQ 2618 part 17.1:2015	$0.12 \mu\text{m} + 0.07 \cdot 10^{-6} \cdot l$	l = measured length

<sup>1)</sup> The expanded uncertainties according to EA-4/02 M:2013 are part of CMC and are the best measurement uncertainties within accreditation. They have a coverage probability of approximately 95 % and have a coverage factor of  $k = 2$  unless stated otherwise. Uncertainties without unit are relative uncertainties referring to the measurement value unless stated otherwise.

**Annex to the accreditation certificate D-K-15118-01-01**
**Permanent Laboratory**
**Calibration and Measurement Capabilities (CMC)**

Measurement quantity / Calibration item	Range	Measurement conditions / procedure	Expanded uncertainty of measurement <sup>1)</sup>	Remarks
Thread gauges single-start cylindrical external and internal threads with straight flanks, symmetrical profile and nominal thread angle 55° bis 60°				
External thread with nominal lead 0.25 mm to 5.5 mm Simple pitch diameter	Nominal diameter: 2 mm to 200 mm	VDI/VDE/DGQ 2618 part 4.8:2006 (option 1) Three wire procedure (vertical to thread axis)	$2.8 \mu\text{m} + 10 \cdot 10^{-6} \cdot d$	$d$ = pitch diameter
Internal thread with nominal lead 0.7 mm to 6.0 mm Simple pitch diameter	Nominal diameter: 4 mm to 200 mm	VDI/VDE/DGQ 2618 part 4.9:2006 (option 1) Two ball procedure (vertical to thread axis)	$2.8 \mu\text{m} + 10 \cdot 10^{-6} \cdot d$	$d$ = Pitch diameter
Angle gauges 90°	Leg length 40 mm to 500 mm	TK 16:2020-01 pointwise measurement	2.4 $\mu\text{m}$	
Inclination measuring instruments elektronic mechanical	$\pm 20$ mm/m	TK 56:2020-01	$2.4 \mu\text{m} + 10 \cdot 10^{-6} \cdot \alpha$	$\alpha$ = Nominal angle
	$\pm 50$ mm/m		21 $\mu\text{m}/\text{m}$	
	2.866° to 45°		0.01°	
	$\pm 90^\circ$		2.3 $\mu\text{m}/\text{m}$	
	Zero point deviation		1.5 $\mu\text{m}/\text{m}$	
Caliper for trailer artifice	to 60 mm	TK 84:2017-03	2 $\mu\text{m}$	
Caliper for trailer artifice	to 120 mm	TK 83:2017-04	8 $\mu\text{m}$	
Layer thickness gauges	20 mm	TK 91:2020-01	$0.7 \mu\text{m} + 180 \cdot 10^{-6} \cdot l$	$l$ = measured length
Calibration foil	20 mm	TK 70:2020-01	$0.8 \mu\text{m} + 130 \cdot 10^{-6} \cdot l$	

<sup>1)</sup> The expanded uncertainties according to EA-4/02 M:2013 are part of CMC and are the best measurement uncertainties within accreditation. They have a coverage probability of approximately 95 % and have a coverage factor of  $k = 2$  unless stated otherwise. Uncertainties without unit are relative uncertainties referring to the measurement value unless stated otherwise.

Annex to the accreditation certificate D-K-15118-01-01

Permanent Laboratory

Calibration and Measurement Capabilities (CMC)

Measurement quantity / Calibration item	Range	Measurement conditions / procedure	Expanded uncertainty of measurement <sup>1)</sup>	Remarks
Gear quantities Profile deviation $F_\alpha$ $f_{i\alpha}$ $f_{H\alpha}$	10 mm $\leq d_b \leq$ 55 mm $L_\alpha \leq$ 4 mm	TK 18.2:2020-02 Substitution measuring with 3D coordinate measuring machines Correction of $F_\alpha$ and $f_{H\alpha}$ by comparison against gear measurement standard with $d_b = 29.88$ mm $L_\alpha = 4$ mm or rather with $d_b = 122.192$ mm $L_\alpha = 24$ mm	1.6 $\mu$ m 0.6 $\mu$ m 1.4 $\mu$ m	Internal and external gears Symbols according to: ISO 1328-1:2013 Evaluation according to: VDI/VDE 2612-1:2018
$F_\alpha$ $f_{i\alpha}$ $f_{H\alpha}$	100 mm $\leq d_b \leq$ 150 mm $L_\alpha \leq$ 24 mm	$d_b = 29.88$ mm $L_\alpha = 4$ mm or rather with $d_b = 122.192$ mm $L_\alpha = 24$ mm	1.6 $\mu$ m 0.6 $\mu$ m 1.4 $\mu$ m	
$F_\alpha$ $f_{i\alpha}$ $f_{H\alpha}$	10 mm $\leq d_b \leq$ 150 mm $L_\alpha \leq$ 24 mm	TK 18.2:2020-02 Measurement with 3D coordinate measuring machines without correction; traceability proved by involute measurement standard with $d_b = 29.88$ mm $L_\alpha = 4$ mm or rather with $d_b = 122.192$ mm $L_\alpha = 24$ mm	2.7 $\mu$ m 1.0 $\mu$ m 1.7 $\mu$ m	
$F_\alpha$ $f_{i\alpha}$ $f_{H\alpha}$	10 mm $\leq d_b \leq$ 500 mm $L_\alpha \leq$ 50 mm	TK 18.1:2020-02 Measurement with 3D coordinate measuring machines without correction; traceability proved by involute measurement standard with $d_b = 29.88$ mm $L_\alpha = 4$ mm or rather with $d_b = 122.192$ mm $L_\alpha = 24$ mm	3.4 $\mu$ m 1.0 $\mu$ m 3.3 $\mu$ m	

<sup>1)</sup> The expanded uncertainties according to EA-4/02 M:2013 are part of CMC and are the best measurement uncertainties within accreditation. They have a coverage probability of approximately 95 % and have a coverage factor of  $k = 2$  unless stated otherwise. Uncertainties without unit are relative uncertainties referring to the measurement value unless stated otherwise.

Annex to the accreditation certificate D-K-15118-01-01

Permanent Laboratory

Calibration and Measurement Capabilities (CMC)

Measurement quantity / Calibration item	Range	Measurement conditions / procedure	Expanded uncertainty of measurement <sup>1)</sup>	Remarks
Helix deviation $F_\beta$ $f_{f\beta}$ $f_{H\beta}$	10 mm $\leq d \leq$ 55 mm $L_\beta \leq$ 40 mm $0^\circ < \beta \leq$ 5°	TK 18.2:2020-02 Substitution measuring with 3D coordinate measuring machines	1.5 $\mu$ m 0.7 $\mu$ m 1.3 $\mu$ m	Internal and external gears Symbols according to: ISO 1328-1:2013 Evaluation according to: VDI/VDE 2612-1:2018
$F_\beta$ $f_{f\beta}$ $f_{H\beta}$	100 mm $\leq d \leq$ 150 mm $L_\beta \leq$ 64 mm $0^\circ < \beta \leq$ 5°	Correction of $F_\beta$ and $f_{H\beta}$ by comparison against gear measurement standard with $d = 34.5$ mm	1.5 $\mu$ m 0.7 $\mu$ m 1.3 $\mu$ m	
$F_\beta$ $f_{f\beta}$ $f_{H\beta}$	100 mm $\leq d \leq$ 150 mm $L_\beta \leq$ 64 mm $10^\circ < \beta \leq$ 20°	$L_\beta = 30$ mm $\beta = 0^\circ$ or rather with	1.6 mm 0.7 mm 1.4 mm	
$F_\beta$ $f_{f\beta}$ $f_{H\beta}$	100 mm $\leq d \leq$ 150 mm $L_\beta \leq$ 40 mm $25^\circ < \beta \leq$ 35°	$d = 104$ mm $L_\beta = 64$ mm $\beta = 0^\circ$ $\beta = 15^\circ$ r+l $\beta = 30^\circ$ r+l	1.9 $\mu$ m 0.7 $\mu$ m 1.7 $\mu$ m	
$F_\beta$ $f_{f\beta}$ $f_{H\beta}$	10 mm $\leq d \leq$ 150 mm $L_\beta \leq$ 40 mm $0^\circ < \beta \leq$ 10°	TK 18.2:2020-02 Measurement with 3D coordinate measuring machines	2.8 $\mu$ m 1.0 $\mu$ m 2.6 $\mu$ m	
$F_\beta$ $f_{f\beta}$ $f_{H\beta}$	100 mm $\leq d \leq$ 150 mm $L_\beta \leq$ 64 mm $5^\circ < \beta \leq$ 10°	without correction; traceability proved by helix measurement standard with	2.8 $\mu$ m 1.0 $\mu$ m 2.6 $\mu$ m	
$F_\beta$ $f_{f\beta}$ $f_{H\beta}$	100 mm $\leq d \leq$ 150 mm $L_\beta \leq$ 64 mm $20^\circ < \beta \leq$ 25°	$d = 34.5$ mm $L_\beta = 30$ mm $\beta = 0^\circ$ or rather with	2.8 $\mu$ m 1.0 $\mu$ m 2.6 $\mu$ m	
$F_\beta$ $f_{f\beta}$ $f_{H\beta}$	100 mm $\leq d \leq$ 150 mm $L_\beta \leq$ 64 mm $35^\circ < \beta \leq$ 45°	$d = 104$ mm $L_\beta = 64$ mm $\beta = 0^\circ$ $\beta = 15^\circ$ r+l $\beta = 30^\circ$ r+l	3.6 $\mu$ m 1.0 $\mu$ m 3.4 $\mu$ m	

<sup>1)</sup> The expanded uncertainties according to EA-4/02 M:2013 are part of CMC and are the best measurement uncertainties within accreditation. They have a coverage probability of approximately 95 % and have a coverage factor of  $k = 2$  unless stated otherwise. Uncertainties without unit are relative uncertainties referring to the measurement value unless stated otherwise.



Annex to the accreditation certificate D-K-15118-01-01

Permanent Laboratory

Calibration and Measurement Capabilities (CMC)

Measurement quantity / Calibration item	Range	Measurement conditions / procedure	Expanded uncertainty of measurement <sup>1)</sup>	Remarks
$F_\beta$ $f_{f\beta}$ $f_{H\beta}$	$10 \text{ mm} \leq d \leq 500 \text{ mm}$ $L_\beta \leq 200 \text{ mm}$ $\beta = 0^\circ$	TK 18.1:2020-02 Measurement with 3D coordinate measuring machines without correction; traceability proved by helix measurement standard with	3.4 $\mu\text{m}$ 1.0 $\mu\text{m}$ 3.2 $\mu\text{m}$	
$F_\beta$ $f_{f\beta}$ $f_{H\beta}$	$10 \text{ mm} \leq d \leq 500 \text{ mm}$ $L_\beta \leq 200 \text{ mm}$ $0^\circ < \beta \leq 35^\circ$	$d = 34.5 \text{ mm}$ $L_\beta = 30 \text{ mm}$ $\beta = 0^\circ$ or rather with $d = 104 \text{ mm}$ $L_\beta = 64 \text{ mm}$ $\beta = 0^\circ$ $\beta = 15^\circ \text{ r+l}$ $\beta = 30^\circ \text{ r+l}$	3.9 $\mu\text{m}$ 1.0 $\mu\text{m}$ 3.7 $\mu\text{m}$	
$F_\beta$ $f_{f\beta}$ $f_{H\beta}$	$10 \text{ mm} \leq d \leq 500 \text{ mm}$ $L_\beta \leq 200 \text{ mm}$ $35^\circ < \beta \leq 45^\circ$	$d = 34.5 \text{ mm}$ $L_\beta = 30 \text{ mm}$ $\beta = 0^\circ$ or rather with $d = 104 \text{ mm}$ $L_\beta = 64 \text{ mm}$ $\beta = 0^\circ$ $\beta = 15^\circ \text{ r+l}$ $\beta = 30^\circ \text{ r+l}$	4.3 $\mu\text{m}$ 1.0 $\mu\text{m}$ 4.2 $\mu\text{m}$	
Pitch deviation $F_p$ $f_p$ $F_x$	$10 \text{ mm} \leq d \leq 500 \text{ mm}$ $\beta = 0^\circ$ $m_n > 0.5 \text{ mm}$	TK 18.2:2020-02 Measurement according to „Rosette method“ with 3D coordinate measuring machines	1.0 $\mu\text{m}$ 0.9 $\mu\text{m}$ 1.1 $\mu\text{m}$	Internal and external gears Symbols according to: ISO 1328-1:2013 Evaluation according to: VDI/VDE 2613:2003
$F_p$ $f_p$ $F_x$	$10 \text{ mm} \leq d \leq 500 \text{ mm}$ $\beta = 0^\circ$ $m_n > 0.5 \text{ mm}$	TK 18.1:2020-02 Measurement with 3D coordinate measuring machines without correction; traceability proved by pitch measurement standard with $d = 67 \text{ mm}$ $m_n = 1 \text{ mm}$	5.1 $\mu\text{m}$ 2.2 $\mu\text{m}$ 5.2 $\mu\text{m}$	
Dimension over balls $M_{dK}$	$10 \text{ mm} \leq M_{dK} \leq 150 \text{ mm}$ $\beta = 0^\circ$ $m_n > 0.5 \text{ mm}$	TK 18.1:2020-02 Measurement of $M_{dK}$ with 3D coordinate measuring machines	$1.4 \mu\text{m} + 11 \cdot 10^{-6} \cdot l$	Internal and external gears Symbols according to: ISO 1328-1:2013

<sup>1)</sup> The expanded uncertainties according to EA-4/02 M:2013 are part of CMC and are the best measurement uncertainties within accreditation. They have a coverage probability of approximately 95 % and have a coverage factor of  $k = 2$  unless stated otherwise. Uncertainties without unit are relative uncertainties referring to the measurement value unless stated otherwise.

Annex to the accreditation certificate D-K-15118-01-01

Permanent Laboratory

Calibration and Measurement Capabilities (CMC)

Measurement quantity / Calibration item	Range	Measurement conditions / procedure	Expanded uncertainty of measurement <sup>1)</sup>	Remarks
<b>Coordinate measuring technology</b> Prismatic, tapered and ball-shaped workpieces	Coordinate measuring machine with calibrated measuring volume of: X = 1200 mm Y = 1000 mm Z = 700 mm	TK 55:2020-02 Tactile measurements with single point probing with a coordinate measuring machine and determination of regular geometries through geometrical parameters (single-points, straight lines, planes, circles, balls, cylinders, tapers, toroid's) using the evaluation software of the coordinate measuring machine. Single-point measuring is carried out with fixed, predefined measuring force. Single point measurements in the form of „Self-centering measurements“ are not used within the accreditation. For ensuring metrological traceability calibration of a similar standard will be realized. Beyond that following limitations should be considered: <ul style="list-style-type: none"> <li>- Measuring points have to be evenly distributed over the form element;</li> <li>- The calibration values can be determined in a multilayer method by averaging in order to reduce the measurement uncertainty.</li> </ul>	The uncertainty of measurement is determined with a uncertainty measurement balance sheet on the basis of the guideline VDI/VDE 2617 part 11:2011. The uncertainty of measurement for specific feedings is specified with a coverage probability of approximately 95 % (coverage factor $k = 2$ ) Exemplary measurement uncertainty for a described measuring tasks: Gauge block with a nominal value of 1000 mm, determined is the expanded uncertainty of the inspection feature „Distance“: $U = 4.8 \mu\text{m}$	For general measuring tasks the measuring uncertainty could be significant differently from the exemplary specified.

<sup>1)</sup> The expanded uncertainties according to EA-4/02 M:2013 are part of CMC and are the best measurement uncertainties within accreditation. They have a coverage probability of approximately 95 % and have a coverage factor of  $k = 2$  unless stated otherwise. Uncertainties without unit are relative uncertainties referring to the measurement value unless stated otherwise.

Annex to the accreditation certificate D-K-15118-01-01

Permanent Laboratory

Calibration and Measurement Capabilities (CMC)

Measurement quantity / Calibration item	Range	Measurement conditions / procedure	Expanded uncertainty of measurement <sup>1)</sup>	Remarks
Radius gauges	1 mm to 2500 mm	TK 86:2020-02	The uncertainty of measurement is determined with a uncertainty measurement balance sheet on the basic of the guideline VDI/VDE 2617 part 11:2011. The uncertainty of measurement is specified with a coverage probability of approximately 95 % (coverage factor $k = 2$ ) Uncertainty of measurement for a measuring problem: Radius with nominal value of 4 mm and an arc of 70°: $U = 10 \mu\text{m}$	
Calibration of control geometries of test and setting gauges with utilities		TK 88:2020-02		
	0 mm to 2000 mm		$38 \mu\text{m} + 26 \cdot 10^{-6} \cdot l$	Calipers, height gauges
	0 mm to 50 mm		$4.9 \mu\text{m} + 2 \cdot 10^{-6} \cdot l$	Micrometers
	0 mm to 1500 mm		$0.2 \mu\text{m} + 4 \cdot 10^{-6} \cdot l$	Horizontal and vertical length measuring device
	15 mm to 150 mm		$2.5 \mu\text{m} + 2.7 \cdot 10^{-6} \cdot l$	Bore gauges
	0° to 360°		0.08°	Universal angle meter, protractors
<b>Force</b> Force measuring devices	0.1 kN to 5 kN	DKD-R 3-3:2018	$3 \cdot 10^{-3}$	
	> 5 kN to 50 kN		$3 \cdot 10^{-3}$	
<b>Torque</b> Manually triggering / indicative operated torque tools	0.1 N·m to 10 N·m	DIN EN ISO 6789-2:2017	$5 \cdot 10^{-3}$	Calibration with calibration device
	10 N·m to 500 N·m		$1 \cdot 10^{-2}$	
Manually triggering / indicative operated torque tools	300 N·m to 3 kN·m	DIN EN ISO 6789-2:2017	$1 \cdot 10^{-2}$	

<sup>1)</sup> The expanded uncertainties according to EA-4/02 M:2013 are part of CMC and are the best measurement uncertainties within accreditation. They have a coverage probability of approximately 95 % and have a coverage factor of  $k = 2$  unless stated otherwise. Uncertainties without unit are relative uncertainties referring to the measurement value unless stated otherwise.

**Annex to the accreditation certificate D-K-15118-01-01**
**Permanent Laboratory**
**Calibration and Measurement Capabilities (CMC)**

Measurement quantity / Calibration item	Range	Measurement conditions / procedure	Expanded uncertainty of measurement <sup>1)</sup>	Remarks
Torque measuring devices, torque measuring chain	0.1 N·m to 10 N·m	DIN 51309:2005	$5 \cdot 10^{-3}$	
Torque wrench calibration devices	10 N·m to 3 kN·m	DKD-R 3-8:2018	$8 \cdot 10^{-3}$	with torque transfer wrench
<b>Pressure</b>				
Absolute pressure $p_{abs}$	0.7 bar to 1.1 bar	DKD-R 6-1:2014 DIN EN 837:1997	1.5 mbar	Pressure medium : gas The measurement uncertainty of the barometer $U_{baro}$ is taken into account
	> 1.1 bar to 61 bar	Method of calibration: $p_{abs} = p_e + p_{amb}$	10 mbar	
Positive gauge pressure $p_e$	0 bar to 60 bar	DKD-R 6-1:2014 DIN EN 837:1997	10 mbar	Pressure medium : gas
Absolute pressure $p_{abs}$	1 bar to 101 bar	DKD-R 6-1:2014 DIN EN 837:1997 Method of calibration: $p_{abs} = p_e + p_{amb}$	$3.1 \cdot 10^{-3} \cdot p_{abs} + 0.01$ bar	Pressure medium: oil $p_{amb}$ = barometric pressure The measurement uncertainty of the barometer $U_{baro}$ is taken into account
	> 101 bar to 1001 bar		$2.9 \cdot 10^{-3} \cdot p_{abs} + 0.07$ bar	
	> 1001 bar to 7001 bar		$2.7 \cdot 10^{-3} \cdot p_{abs} + 0.8$ bar	
Positive gauge pressure $p_e$	0 bar to 100 bar	DKD-R 6-1:2014 DIN EN 837:1997	$3.1 \cdot 10^{-3} \cdot p_e + 0.01$ bar	Pressure medium: oil
	> 100 bar to 1000 bar		$2.9 \cdot 10^{-3} \cdot p_e + 0.07$ bar	
	> 1000 bar to 7000 bar		$2.7 \cdot 10^{-3} \cdot p_e + 0.8$ bar	
<b>Weighing instruments</b>				
Nonautomatic weighing instruments	to 30 kg	EURAMET cg 18 version 4.0	$1.3 \cdot 10^{-6}$	with weights OIML R 111-1:2004 according to the class E2
	to 100 kg		$6.6 \cdot 10^{-5}$	with weights OIML R 111-1:2004 according to the class M1
<b>Temperature quantities</b>				
Resistance thermometers with display unit	0 °C to 200 °C	DKD-R 5-1:2018	0.25 K	
	> 200 °C to 400 °C		0.4 K	
Thermocouples with display unit	0 °C to 200 °C	DKD-R 5-3:2018	0.4 K	
	> 200 °C to 400 °C		0.5 K	

<sup>1)</sup> The expanded uncertainties according to EA-4/02 M:2013 are part of CMC and are the best measurement uncertainties within accreditation. They have a coverage probability of approximately 95 % and have a coverage factor of  $k = 2$  unless stated otherwise. Uncertainties without unit are relative uncertainties referring to the measurement value unless stated otherwise.

Annex to the accreditation certificate D-K-15118-01-01

**On-site Calibration**

Calibration and Measurement Capabilities (CMC)

Measurement quantity / Calibration item	Range	Measurement conditions / procedure	Expanded uncertainty of measurement <sup>1)</sup>	Remarks
<b>Length</b>				
Surface plates Flatness deviation	to 50 µm	VDI/VDE/DGQ 2618 part 6.2:2014 to 8 m edge length	$1.3 \mu\text{m} + 1.3 \cdot 10^{-6} \cdot l$	$l$ = measured length with inclination measuring instruments
Height gauges	0 mm to 1000 mm	VDI/VDE/DGQ 2618 part 16.1:2009	$1.7 \mu\text{m} + 1.2 \cdot 10^{-6} \cdot l$	$l$ = measured length
Height gauges	0 mm to 1000 mm	VDI/VDE/DGQ 2618 part 16.1:2009	0.9 µm	till 1000 mm lead length
Deviation from straightness and perpendicularity	to 40 µm		4 µm	
Horizontal length measuring devices	0 mm to 5000 mm	VDI/VDE/DGQ 2618 part 17.1:2015	$0.12 \mu\text{m} + 0.07 \cdot 10^{-6} \cdot l$	$l$ = measured length
Height calipers with analogue display	0 mm to 600 mm	VDI/VDE/DGQ 2618 part 9.3:2006	$30 \mu\text{m} + 30 \cdot 10^{-6} \cdot l$	
with digital display			$20 \mu\text{m} + 30 \cdot 10^{-6} \cdot l$	
<b>Coordinate measuring technology</b>				
Measuring projectors Measuring microscopes	Devices featuring a measuring plane with a face diagonal ≤ 450 mm	Calibration of metrological characteristics according to guideline DKD-R 4-3 part 18.1:2018, and the following standards and guidelines DIN EN ISO 10360 VDI/VDE 2617		
		Determination of probing error <i>PS-ID(OT)</i> with a graduated scale made of glass according to VDI/VDE 2617 part 6.1:2007	0.8 µm	
		The error of indication for size measurement <i>E-ID(OT)</i> is determined with a graduated scale made of glass according to VDI/VDE 2617 part 6.1:2007	$1.6 \mu\text{m} + 1 \cdot 10^{-6} \cdot l$	$l$ = measured length
<b>Torque</b>				
Torque wrench calibration devices	10 N·m to 3 kN·m	DKD-R 3-8:2018	$8 \cdot 10^{-3}$	with torque transfer wrench
Manually triggering / indicative operated torque tools	10 N·m to 500 N·m	DIN EN ISO 6789-2:2017	$1 \cdot 10^{-2}$	Calibration with calibration device
	300 N·m to 3 kN·m		$1 \cdot 10^{-2}$	

<sup>1)</sup> The expanded uncertainties according to EA-4/02 M:2013 are part of CMC and are the best measurement uncertainties within accreditation. They have a coverage probability of approximately 95 % and have a coverage factor of  $k = 2$  unless stated otherwise. Uncertainties without unit are relative uncertainties referring to the measurement value unless stated otherwise.

Annex to the accreditation certificate D-K-15118-01-01

**On-site Calibration**

Calibration and Measurement Capabilities (CMC)				
Measurement quantity / Calibration item	Range	Measurement conditions / procedure	Expanded uncertainty of measurement <sup>1)</sup>	Remarks
<b>Druck</b> Absolute pressure $p_{abs}$	0.7 bar to 1.1 bar	DKD-R 6-1:2014 DIN EN 837:1997 Method of calibration: $p_{abs} = p_e + p_{amb}$	1.5 mbar	Pressure medium: gas The measurement uncertainty of the barometer $U_{baro}$ is taken into account
Positive gauge pressure $p_e$	0 bar to 60 bar	DKD-R 6-1:2014 DIN EN 837:1997	10 mbar	Pressure medium: gas
Absolute pressure $p_{abs}$	1 bar to 101 bar	DKD-R 6-1:2014 DIN EN 837:1997 Method of calibration: $p_{abs} = p_e + p_{amb}$	$3.1 \cdot 10^{-3} \cdot p_{abs} + 0.01$ bar	Pressure medium: oil $p_{amb}$ = atmosphärischer Luftdruck The measurement uncertainty of the barometer $U_{baro}$ is taken into account
	> 101 bar to 1001 bar		$2.9 \cdot 10^{-3} \cdot p_{abs} + 0.07$ bar	
	> 1001 bar to 7001 bar		$2.7 \cdot 10^{-3} \cdot p_{abs} + 0.8$ bar	
Positive gauge pressure $p_e$	0 bar to 100 bar	DKD-R 6-1:2014 DIN EN 837:1997	$3.1 \cdot 10^{-3} \cdot p_e + 0.01$ bar	Pressure medium: oil
	> 100 bar to 1000 bar		$2.9 \cdot 10^{-3} \cdot p_e + 0.07$ bar	
	> 1000 bar to 7000 bar		$2.7 \cdot 10^{-3} \cdot p_e + 0.8$ bar	
<b>Weighing instruments</b> Nonautomatic weighing instruments	to 30 kg	EURAMET cg 18 Version 4.0	$1.3 \cdot 10^{-6}$	with weights OIML R 111-1:2004 according to the class E2
	to 100 kg		$6.6 \cdot 10^{-5}$	with weights OIML R 111-1:2004 according to the class M1

**Mobile Laboratory**

Calibration and Measurement Capabilities (CMC)				
Measurement quantity / Calibration item	Range	Measurement conditions / procedure	Expanded uncertainty of measurement <sup>1)</sup>	Remarks
<b>Länge</b> Gap gauges	5 mm to 200 mm	VDI/VDE/DGQ 2618 part 4.7:2005	$3 \mu\text{m} + 10 \cdot 10^{-6} \cdot l$	$l$ = measured length
Calipers for external, internal and depth dimensions	0 mm to 300 mm	VDI/VDE/DGQ 2618 part 9.1:2006	$30 \mu\text{m} + 30 \cdot 10^{-6} \cdot l$	
Depth calipers	0 mm to 300 mm	VDI/VDE/DGQ 2618 part 9.2:2006	$30 \mu\text{m} + 30 \cdot 10^{-6} \cdot l$	

<sup>1)</sup> The expanded uncertainties according to EA-4/02 M:2013 are part of CMC and are the best measurement uncertainties within accreditation. They have a coverage probability of approximately 95 % and have a coverage factor of  $k = 2$  unless stated otherwise. Uncertainties without unit are relative uncertainties referring to the measurement value unless stated otherwise.

Date of issue: 22.10.2020

Valid from: 22.10.2020

Annex to the accreditation certificate D-K-15118-01-01

**Mobile Laboratory**

Calibration and Measurement Capabilities (CMC)

Measurement quantity / Calibration item	Range	Measurement conditions / procedure	Expanded uncertainty of measurement <sup>1)</sup>	Remarks
Height calipers with analogue display	0 mm to 600 mm	VDI/VDE/DGQ 2618 part 9.3:2006	$30 \mu\text{m} + 30 \cdot 10^{-6} \cdot l$	$l = \text{measured length}$
			$20 \mu\text{m} + 30 \cdot 10^{-6} \cdot l$	
Micrometers	0 mm to 100 mm	VDI/VDE/DGQ 2618 part 10.1:2001	$3 \mu\text{m} + 10 \cdot 10^{-6} \cdot l$	$l = \text{final value of the measuring range}$
	> 100 mm to 500 mm		$4 \mu\text{m} + 10 \cdot 10^{-6} \cdot l$	
Internal micrometers with three-point contact	3 mm to 200 mm	VDI/VDE/DGQ 2618 part 10.8:2002	$4 \mu\text{m} + 10 \cdot 10^{-6} \cdot d$	$d = \text{measured diameter}$
Internal measuring instruments	3 mm to 200 mm	TK 57:2017-03	$2 \mu\text{m} + 10 \cdot 10^{-6} \cdot d$	
Dial gauges Scale interval > 1 $\mu\text{m}$	to 100 mm	VDI/VDE/DGQ 2618 part 11.1:2014	$3 \mu\text{m} + 10 \cdot 10^{-6} \cdot l$	$l = \text{measured length}$
			$2 \mu\text{m} + 10 \cdot 10^{-6} \cdot l$	
Dial indicators	to 3 mm	VDI/VDE/DGQ 2618 part 11.2:2002	1.1 $\mu\text{m}$	
Lever gauges	to 1.6 mm	VDI/VDE/DGQ 2618 part 11.3:2002	1.2 $\mu\text{m}$	
Thickness gauges Scale interval 1 $\mu\text{m}$	0 mm to 30 mm	VDI/VDE/DGQ 2618 part 12.1:2005	$1.3 \mu\text{m} + 10 \cdot 10^{-6} \cdot l$	$l = \text{measured length}$
			$6 \mu\text{m} + 10 \cdot 10^{-6} \cdot l$	
Feeler gauges	10 $\mu\text{m}$ to 2 mm	DIN 2275:2014	1 $\mu\text{m}$	
Lever gauges (quicktests) for external measurements	0 mm to 100 mm	VDI/VDE/DGQ 2618 part 12.1:2005	6 $\mu\text{m}$	
Lever gauges (quicktests) for internal measurements	2.5 mm to 500 mm	VDI/VDE/DGQ 2618 part 13.1:2005	6 $\mu\text{m}$	
Internal micrometers with two-point contact	25 mm to 100 mm	VDI/VDE/DGQ 2618 part 10.7:2010	$3 \mu\text{m} + 10 \cdot 10^{-6} \cdot l$	
	> 100 mm to 500 mm		$4 \mu\text{m} + 10 \cdot 10^{-6} \cdot l$	
	> 500 mm to 800 mm		$5 \mu\text{m} + 10 \cdot 10^{-6} \cdot l$	
Extensions for internal micrometers with two-point contact	25 mm to 500 mm	VDI/VDE/DGQ 2618 part 10.7:2010	$2 \mu\text{m} + 5 \cdot 10^{-6} \cdot l$	
	> 500 mm to 800 mm		$3.5 \mu\text{m} + 5 \cdot 10^{-6} \cdot l$	
Micrometers with interchangeable inserts	0 mm to 100 mm	VDI/VDE/DGQ 2618 part 10.2:2010	$3 \mu\text{m} + 10 \cdot 10^{-6} \cdot l$	
	> 100 mm to 300 mm		$5 \mu\text{m} + 10 \cdot 10^{-6} \cdot l$	

<sup>1)</sup> The expanded uncertainties according to EA-4/02 M:2013 are part of CMC and are the best measurement uncertainties within accreditation. They have a coverage probability of approximately 95 % and have a coverage factor of  $k = 2$  unless stated otherwise. Uncertainties without unit are relative uncertainties referring to the measurement value unless stated otherwise.

Annex to the accreditation certificate D-K-15118-01-01

Mobile Laboratory

Calibration and Measurement Capabilities (CMC)

Measurement quantity / Calibration item	Range	Measurement conditions / procedure	Expanded uncertainty of measurement <sup>1)</sup>	Remarks
Micrometers with dial indicators	0 mm to 100 mm	VDI/VDE/DGQ 2618 part 10.3:2002	$3 \mu\text{m} + 10 \cdot 10^{-6} \cdot l$	$l$ = measured length
Micrometers heads	0 mm to 100 mm	VDI/VDE/DGQ 2618 part 10.4:2008	$3 \mu\text{m} + 10 \cdot 10^{-6} \cdot l$	
	> 100 mm to 500 mm		$4 \mu\text{m} + 10 \cdot 10^{-6} \cdot l$	
	> 500 mm to 800 mm		$5 \mu\text{m} + 10 \cdot 10^{-6} \cdot l$	
Depth micrometers	0 mm to 100 mm	VDI/VDE/DGQ 2618 part 10.5:2010	$3 \mu\text{m} + 10 \cdot 10^{-6} \cdot l$	
	> 100 mm to 500 mm		$4 \mu\text{m} + 10 \cdot 10^{-6} \cdot l$	
	> 500 mm to 800 mm		$5 \mu\text{m} + 10 \cdot 10^{-6} \cdot l$	
Reference gauges for micrometers	25 mm to 800 mm	VDI/VDE/DGQ 2618 part 4.4:2009	$3 \mu\text{m} + 5 \cdot 10^{-6} \cdot l$	
Protractors	0° to 360°	VDI/VDE/DGQ 2618 part 7.2:2008		
Flatness deviation			5 μm	
Parallelism deviation			5 μm	
Angle				
Scale interval 5'	0° to 360°		4'	
Scale interval 1°	0° to 180°		24'	
<b>Druck</b>				
Absolute pressure $p_{\text{abs}}$	0.7 bar to 1.1 bar	DKD-R 6-1:2014 DIN EN 837:1997 Method of calibration: $p_{\text{abs}} = p_e + p_{\text{amb}}$	1.5 mbar	Pressure medium : gas
	> 1.1 bar to 61 bar		10 mbar	The measurement uncertainty of the barometer $U_{\text{baro}}$ is taken into account
Positive gauge pressure $p_e$	0 bar to 60 bar	DKD-R 6-1:2014 DIN EN 837:1997	10 mbar	Pressure medium : gas
Absolute pressure $p_{\text{abs}}$	1 bar to 101 bar	DKD-R 6-1:2014 DIN EN 837:1997 Method of calibration: $p_{\text{abs}} = p_e + p_{\text{amb}}$	$3.1 \cdot 10^{-3} \cdot p_{\text{abs}} + 0.01 \text{ bar}$	Pressure medium: oil
	> 101 bar to 1001 bar		$2.9 \cdot 10^{-3} \cdot p_{\text{abs}} + 0.07 \text{ bar}$	$p_{\text{amb}}$ = barometric pressure
	> 1001 bar to 7001 bar		$2.7 \cdot 10^{-3} \cdot p_{\text{abs}} + 0.8 \text{ bar}$	The measurement uncertainty of the barometer $U_{\text{baro}}$ is taken into account
Positive gauge pressure $p_e$	0 bar to 100 bar	DKD-R 6-1:2014 DIN EN 837:1997	$3.1 \cdot 10^{-3} \cdot p_e + 0.01 \text{ bar}$	Pressure medium: oil
	> 100 bar to 1000 bar		$2.9 \cdot 10^{-3} \cdot p_e + 0.07 \text{ bar}$	
	> 1000 bar to 7000 bar		$2.7 \cdot 10^{-3} \cdot p_e + 0.8 \text{ bar}$	

<sup>1)</sup> The expanded uncertainties according to EA-4/02 M:2013 are part of CMC and are the best measurement uncertainties within accreditation. They have a coverage probability of approximately 95 % and have a coverage factor of  $k = 2$  unless stated otherwise. Uncertainties without unit are relative uncertainties referring to the measurement value unless stated otherwise.



**Annex to the accreditation certificate D-K-15118-01-01**

**Abbreviations used:**

CMC	Calibration and measurement capabilities
DIN	Deutsches Institut für Normung e.V.
DKD-R	Guideline of Deutscher Kalibrierdienst (DKD), published by Physikalisch-Technische Bundesanstalt
DGQ	Deutsche Gesellschaft für Qualität e.V.
EURAMET	European Association of National Metrology Institutes
TK	Calibration Guide of Kessler-QMP GmbH
VDE	Verband der Elektrotechnik, Elektronik und Informationstechnik e.V.
VDI	Verein Deutscher Ingenieure e.V.

$\beta$	Helix angle	$f_{f\beta}$	Helix form deviation
$d$	Reference diameter	$f_{H\beta}$	Helix slope deviation
$d_b$	Base diameter	$F_p$	Cumulative pitch deviation
$F_\alpha$	Total profile deviation	$f_p$	Single pitch deviation
$f_{H\alpha}$	Profile slope deviation	$L_\alpha$	Profile evaluation range
$f_{f\alpha}$	Profile form deviation	$L_\beta$	Helix evaluation range
$F_\beta$	Total helix deviation	$m_n$	Normal module

<sup>1)</sup> The expanded uncertainties according to EA-4/02 M:2013 are part of CMC and are the best measurement uncertainties within accreditation. They have a coverage probability of approximately 95 % and have a coverage factor of  $k = 2$  unless stated otherwise. Uncertainties without unit are relative uncertainties referring to the measurement value unless stated otherwise.