

Deutsche Akkreditierungsstelle GmbH

Annex to the Accreditation Certificate D-K-15199-01-00 according to DIN EN ISO/IEC 17025:2005

Period of validity: 15.09.2016 to 14.09.2021

Date of issue: 15.09.2016

Holder of certificate:

**Frenco GmbH, Verzahnungstechnik, Messtechnik
Jakob-Baier-Straße 3, 90518 Altdorf**

Head: Dipl.-Ing. (FH) Jan Kühl
Deputy: Jürgen Stellwag

Accredited as calibration laboratory since: 17.04.2000

Calibrations in the fields:

Dimensional quantities

Length

- **Gear quantities**

Abbreviations used: see last page

Annex to the accreditation certificate D-K-15199-01-00

Permanent Laboratory

Measurement quantity / Calibration item	Range	Measurement conditions / procedure	Best measurement capability ¹⁾	Remarks
Length				
Gear quantities	Base diameter: d_b	Substitution measuring with gear devices:		External gears
Involute artefact	Evaluation range: L_α	Correction of $F_\alpha, f_{H\alpha}$ by comparison against involute artefact with	1.8 μm 1.4 μm 1.0 μm	Symbols according to ISO 1328-1:2013
F_α	$30 \text{ mm} \leq d_b \leq 50 \text{ mm}$			
$f_{H\alpha}$	$10 \text{ mm} \leq L_\alpha \leq 16 \text{ mm}$			
$f_{i\alpha}$				
F_α	$25 \text{ mm} \leq d_b \leq 60 \text{ mm}$	$d_b = 40 \text{ mm}$	2.0 μm	Evaluation according to guidelines
$f_{H\alpha}$	$7 \text{ mm} \leq L_\alpha \leq 18 \text{ mm}$	$L_\alpha = 13 \text{ mm}$	1.8 μm	
$f_{i\alpha}$			1.0 μm	
F_α	$80 \text{ mm} \leq d_b \leq 120 \text{ mm}$	Correction of $F_\alpha, f_{i\alpha}$ by comparison against involute artefact with	1.4 μm	VDI/VDE 2607:2000 VDI/VDE 2612:2000
$f_{H\alpha}$	$14 \text{ mm} \leq L_\alpha \leq 42 \text{ mm}$		1.0 μm	
$f_{i\alpha}$			0.9 μm	
F_α	$60 \text{ mm} \leq d_b \leq 130 \text{ mm}$	$d_b = 93.96 \text{ mm}$	1.7 μm	
$f_{H\alpha}$	$8 \text{ mm} \leq L_\alpha \leq 48 \text{ mm}$	$L_\alpha = 37 \text{ mm}$	1.4 μm	
$f_{i\alpha}$		or with $d_b = 100 \text{ mm}$	0.9 μm	
F_α	$60 \text{ mm} \leq d_b \leq 130 \text{ mm}$	$L_\alpha = 37 \text{ mm}$	1.8 μm	
$f_{H\alpha}$	$20 \text{ mm} \leq L_\alpha \leq 53 \text{ mm}$		1.5 μm	
$f_{i\alpha}$			1.0 μm	
F_α	$8 \text{ mm} \leq d_b \leq 150 \text{ mm}$	Measurement without correction; traceability proved by involute artefact with	2.1 μm	
$f_{H\alpha}$	$L_\alpha \leq 74 \text{ mm}$	$d_b = 100 \text{ mm}, L_\alpha = 37 \text{ mm}$	1.8 μm	
$f_{i\alpha}$		or with $d_b = 93.96 \text{ mm}$	0.9 μm	
		$L_\alpha = 37 \text{ mm}$		

¹⁾ The best measurement capabilities are stated according to DAKKS-DKD-3 (EA-4/02). These are expanded uncertainties of measurement with a coverage probability of 95% and have a coverage factor of $k=2$ unless stated otherwise. Uncertainties without unit are relative uncertainties referring to the measurement value unless stated otherwise.

Annex to the accreditation certificate D-K-15199-01-00

Measurement quantity / Calibration item	Range	Measurement conditions / procedure	Best measurement capability ¹⁾	Remarks
Gear quantities Helix artefacts	Reference diameter: d Helix angle: β Evaluation range: L_β	Substitution measuring with gear devices:		
F_β $f_{H\beta}$ f_{Φ}	$28 \text{ mm} \leq d \leq 36 \text{ mm}$ $10^\circ \leq \beta \leq 20^\circ$ $34 \text{ mm} \leq L_\beta \leq 46 \text{ mm}$	Correction of $F_\beta, f_{H\beta}$ by comparison against helix artefact with	$1.9 \mu\text{m}$ $1.6 \mu\text{m}$ $1.0 \mu\text{m}$	External gears Symbols according to ISO 1328-1:2013
F_β $f_{H\beta}$ f_{Φ}	$20 \text{ mm} \leq d \leq 45 \text{ mm}$ $10^\circ \leq \beta \leq 20^\circ$ $30 \text{ mm} \leq L_\beta \leq 50 \text{ mm}$	$d = 32 \text{ mm}$ $\beta = 15^\circ \text{ r+l}$ $L_\beta = 40 \text{ mm}$	$2.0 \mu\text{m}$ $1.8 \mu\text{m}$ $1.0 \mu\text{m}$	Evaluation according to guidelines
F_β $f_{H\beta}$ f_{Φ}	$85 \text{ mm} \leq d \leq 125 \text{ mm}$ $\beta = 0^\circ$ $56 \text{ mm} \leq L_\beta \leq 102 \text{ mm}$	Correction of $F_\beta, f_{H\beta}$ by comparison against helix artefact with	$1.3 \mu\text{m}$ $0.9 \mu\text{m}$ $0.9 \mu\text{m}$	VDI/VDE 2607:2000 VDI/VDE 2612:2000
F_β $f_{H\beta}$ f_{Φ}	$70 \text{ mm} \leq d \leq 135 \text{ mm}$ $\beta = 0^\circ$ $30 \text{ mm} \leq L_\beta \leq 120 \text{ mm}$	$d = 100 \text{ mm}$ $\beta = 0^\circ$ $\beta = 15^\circ \text{ r+l}$ $\beta = 30^\circ \text{ r+l}$ $L_\beta = 94 \text{ mm}$	$1.5 \mu\text{m}$ $1.1 \mu\text{m}$ $0.9 \mu\text{m}$	
F_β $f_{H\beta}$ f_{Φ}	$85 \text{ mm} \leq d \leq 125 \text{ mm}$ $10^\circ \leq \beta \leq 20^\circ$ $56 \text{ mm} \leq L_\beta \leq 102 \text{ mm}$		$1.4 \mu\text{m}$ $1.0 \mu\text{m}$ $0.9 \mu\text{m}$	
F_β $f_{H\beta}$ f_{Φ}	$70 \text{ mm} \leq d \leq 135 \text{ mm}$ $7^\circ \leq \beta \leq 23^\circ$ $46 \text{ mm} \leq L_\beta \leq 112 \text{ mm}$		$1.6 \mu\text{m}$ $1.3 \mu\text{m}$ $0.9 \mu\text{m}$	
F_β $f_{H\beta}$ f_{Φ}	$85 \text{ mm} \leq d \leq 125 \text{ mm}$ $25^\circ \leq \beta \leq 35^\circ$ $56 \text{ mm} \leq L_\beta \leq 102 \text{ mm}$		$1.5 \mu\text{m}$ $1.2 \mu\text{m}$ $0.9 \mu\text{m}$	
F_β $f_{H\beta}$ f_{Φ}	$70 \text{ mm} \leq d \leq 135 \text{ mm}$ $23^\circ \leq \beta \leq 37^\circ$ $46 \text{ mm} \leq L_\beta \leq 112 \text{ mm}$		$1.7 \mu\text{m}$ $1.4 \mu\text{m}$ $0.9 \mu\text{m}$	
F_β $f_{H\beta}$ f_{Φ}	$10 \text{ mm} \leq d \leq 160 \text{ mm}$ $\beta = 0^\circ$ $10 \text{ mm} \leq L_\beta \leq 130 \text{ mm}$	Measurement without correction; traceability proved by involute artefact with	$1.5 \mu\text{m}$ $1.2 \mu\text{m}$ $0.9 \mu\text{m}$	
F_β $f_{H\beta}$ f_{Φ}	$10 \text{ mm} \leq d \leq 160 \text{ mm}$ $0^\circ \leq \beta \leq 20^\circ$ $10 \text{ mm} \leq L_\beta \leq 130 \text{ mm}$	$d = 100 \text{ mm}, L_\beta = 94 \text{ mm}$ $\beta = 0^\circ$ $\beta = 15^\circ \text{ r+l}$ $\beta = 30^\circ \text{ r+l}$	$1.7 \mu\text{m}$ $1.4 \mu\text{m}$ $0.9 \mu\text{m}$	
F_β $f_{H\beta}$ f_{Φ}	$10 \text{ mm} \leq d \leq 160 \text{ mm}$ $20^\circ \leq \beta \leq 40^\circ$ $10 \text{ mm} \leq L_\beta \leq 130 \text{ mm}$		$2.0 \mu\text{m}$ $1.7 \mu\text{m}$ $0.9 \mu\text{m}$	

¹⁾ The best measurement capabilities are stated according to DAKKS-DKD-3 (EA-4/02). These are expanded uncertainties of measurement with a coverage probability of 95% and have a coverage factor of $k=2$ unless stated otherwise. Uncertainties without unit are relative uncertainties referring to the measurement value unless stated otherwise.

Annex to the accreditation certificate D-K-15199-01-00

Measurement quantity / Calibration item	Range	Measurement conditions / procedure	Best measurement capability ¹⁾	Remarks
Pitch and runout F_P f_P F_r	Reference diameter: d Normal module: M_n $5 \text{ mm} \leq d \leq 350 \text{ mm}$ $M_n \geq 0.5$	According to „Rosette method“ on gear measuring device. To be carried out according to working instructions of the Laboratory QM	0.7 μm 0.6 μm 1.0 μm	External gears Symbols according to ISO 1328-1:2013 Evaluation according to guidelines VDI/VDE 2613:2003
Dimension over balls M_{dK}	Dimension over balls: M_{dK} Helix angle: β Normal module: M_n $M_{dK} \leq 240 \text{ mm}$ $\beta = 0^\circ$ $M_n \geq 0.5$	Measurement of M_{dK} on length comparator compared to traceable setting standard in accordance with working instructions of the Laboratory QM	1.2 μm	
M_{dK}	$M_{dK} \leq 240 \text{ mm}$ $\beta \geq 0^\circ$ $M_n \geq 0.5$		1.2 μm	

Abbreviations used:

VDI/VDE 2607/2612/2613 VDI-Guideline on gearing

β	Helix angle	F_p	Total pitch error
d	Reference diameter	f_p	Single pitch deviation
d_b	Base diameter	F_r	Runout error
F_α	Total profile deviation	L_α	Profile evaluation range
$f_{f\alpha}$	Profile form deviation	L_β	Helix evaluation range
F_β	Total helix deviation	M_{dK}	Dimension over balls
$f_{f\beta}$	Helix form deviation	M_n	Normal module
$f_{H\beta}$	Helix slope deviation	r+l	Right hand and left hand

¹⁾ The best measurement capabilities are stated according to DAKKS-DKD-3 (EA-4/02). These are expanded uncertainties of measurement with a coverage probability of 95% and have a coverage factor of $k=2$ unless stated otherwise. Uncertainties without unit are relative uncertainties referring to the measurement value unless stated otherwise.