

# Deutsche Akkreditierungsstelle GmbH

## Annex to the Accreditation Certificate D-K-17060-01-00 according to DIN EN ISO/IEC 17025:2018

**Valid from: 13.02.2020**

Date of issue: 13.02.2020

Holder of certificate:

**ifm prover gmbh Sensorik für die Prozeß- und Verfahrenstechnik  
Waldesch 9, 88069 Tettngang**

Calibration in the fields:

**Thermodynamic quantities**

**Temperature quantities**

- Resistance thermometers
- Direct reading thermometers
- Temperature transmitters, data loggers

**Mechanical quantities**

- Pressure

**Fluid quantities**

- Gas flow rate

**Within the scope of accreditation marked with \*) the calibration laboratory is permitted, without being required to inform and obtain prior approval from DAkkS, to use calibration standards or equivalent calibration procedures listed here with different issue dates.**

**The calibration laboratory maintains a current list of all calibration standards / equivalent calibration procedures within the flexible scope of accreditation.**

Abbreviations used: see last page

**Annex to the accreditation certificate D-K-17060-01-00**

**Permanent Laboratory**

**Calibration and Measurement Capabilities (CMC)**

Measurement quantity / Calibration item	Range	Measurement conditions / procedure	Expanded uncertainty of measurement <sup>1)</sup>	Remarks
<b>Temperature</b> Resistance thermometers also direct reading or with measuring device or transmitter <sup>*)</sup>	-20 °C to 150 °C	circulated liquid bath, DKD-R 5-1:2018	0.1 K	Comparison with standard resistance thermometers in thermostatic baths
<b>Pressure</b> Negative and positive gauge pressure $p_e^{*)}$	-1 bar to -0.03 bar 0 bar; 0.015 bar to 1 bar	DKD-R 6-1:2014	$2 \cdot 10^{-4} \cdot p_e$ ; but not less than 20 $\mu$ bar	Pressure medium: Gas
Positive gauge pressure $p_e^{*)}$	0 bar; 1 bar to 35 bar			
	> 35 bar to 700 bar		$2 \cdot 10^{-4} \cdot p_e$ ; but not less than 15 mbar	
<b>Gas flow rate</b> Volume flow rate of flowing gases (standard volume flow)	0.05 m <sup>3</sup> /h to < 1.5 m <sup>3</sup> /h 1.5 m <sup>3</sup> /h to 1000 m <sup>3</sup> /h	KV-F001:2017-06	1.0 %	Calibration gas: dry air below 2.7 bar (absolute pressure).
			0.60 %	
Mass flow rate of flowing gases	0.06 kg/h to < 1.9 kg/h 1.9 kg/h to 1200 kg/h		1.0 %	Devices with visual indicator and / or analogue output (current, voltage)
			0.60 %	

**Abbreviations used:**

CMC	Calibration and measurement capabilities
DKD-R	Guideline of Deutscher Kalibrierdienst (DKD), published by Physikalisch-Technische Bundesanstalt
KV-F001	Procedure of ifm prover gmbh

<sup>1)</sup> The expanded uncertainties according to EA-4/02 M:2013 are part of CMC and are the best measurement uncertainties within accreditation. They have a coverage probability of approximately 95 % and have a coverage factor of  $k = 2$  unless stated otherwise. Uncertainties without unit are relative uncertainties referring to the measurement value unless stated otherwise.