

Deutsche Akkreditierungsstelle GmbH

Annex to the Accreditation Certificate D-K-17601-01-00 according to DIN EN ISO/IEC 17025:2018

Valid from: 15.12.2020

Date of issue 08.04.2021

Holder of certificate:

**H+H High Voltage Technology GmbH
Im Kurzen Busch 15, 58640 Iserlohn**

Calibration in the fields:

Electrical quantities

DC and low frequency quantities

- DC voltage ^{*)}
- AC voltage ^{*)}
- High voltage quantities ^{*)}
- High voltage impulse quantities ^{*)}
- Charge

High frequency and radiation quantities

High frequency quantities

- Rise time
- Waveform quantities

^{*)} also on site calibration

Within the scope of accreditation marked with ^{**)}, the calibration laboratory is permitted, without being required to inform and obtain prior approval from DAkkS, to use calibration standards or equivalent calibration procedures listed here with different issue dates. The calibration laboratory maintains a current list of all calibration standards / equivalent calibration procedures within the flexible scope of accreditation.

The management system requirements in DIN EN ISO/IEC 17025 are written in language relevant to operations of calibration laboratories and operate generally in accordance with the principles of DIN EN ISO 9001.

*The certificate together with its annex reflects the status at the time of the date of issue. The current status of the scope of accreditation can be found in the database of accredited bodies of Deutsche Akkreditierungsstelle GmbH.
<https://www.dakks.de/en/content/accredited-bodies-dakks>*

Annex to the accreditation certificate D-K-17601-01-00

Permanent Laboratory

Calibration and Measurement Capabilities (CMC)

Measurement quantity / Calibration item	Range	Measurement conditions / procedure	Expanded uncertainty of measurement ¹⁾	Remarks	
DC voltage	-1 kV to -100 kV		0.7 %		
	1 kV to 100 kV		0.7 %		
AC voltage	1 kV to 100 kV	50 Hz	0.4 %		
	50 kV to 300 kV	50 Hz	0.8 %		
Burst voltage (LI) Amplitude	-10 kV to -500 kV		0.8 %	LI = Lightning Impulse	
	10 kV to 500 kV		0.8 %		
Time parameter T_1 T_2	0.8 μ s to 1.6 μ s		2.2 %	T_1 = Front slope time T_2 = Back slope-Half lifetime	
	40 μ s to 60 μ s		2.2 %		
Impulse charge ^{**)}	1 pC to 100 nC	DIN EN 60270-2:2011-10	0.03 · q + 0.5 pC	q: charge quantity	
Rise time t_r	3 ns to 1,0 μ s		4,1 %		
Electrostatic discharge (ESD) ^{**)} Current pulse I_p	1 A to 35 A	DIN EN 61000-4-2:2009-12	3.1 %	I_p = first burst current peak I_{30} = current at 30 ns I_{60} = current at 60 ns U_L = free state-voltage	
	Basic values				
	Current pulse I_{30}		1 A to 35 A		1.9 %
	Current pulse I_{60}		1 A to 35 A		1.9 %
	Rise time t_r		0.6 ns to 1 ns		3.9 %
DC voltage U_L	1 kV to 100 kV		0.7 %		
EFT/B Burst ^{**)} Voltage pulse	100 V to 2200 V	DIN EN 61000-4-4:2013-04	1.7 %	R_L = load resistor on R_L = 50 Ω on R_L = 1000 Ω	
	100 V to 4400 V		2.1 %		
	Rise time and Impulse width		3 ns to 1 μ s		4.1 %
	Burst duration and burst periode		100 ns to 1 s		2.0 %
Surge ^{**)} DC-amplitude	500 V to 12 kV	DIN EN 61000-4-5:2015-03	1.9 %		
	Current amplitude		1 A to 120 kA	3.0 %	
	Rise time and Impulse width		0.8 μ s to 1.0 ms	1.0 %	
Dips Voltage amplitude ^{**)}	1 V to 700 V	DIN EN 61000-4-11:2005-02 16 $\frac{1}{2}$ Hz ; 50 Hz ; 60 Hz	1.3 %		
	Time interval		100 ns to 1 s	0.6 %	

¹⁾ The expanded uncertainties according to EA-4/02 M:2013 are part of CMC and are the best measurement uncertainties within accreditation. They have a coverage probability of approximately 95 % and have a coverage factor of $k = 2$ unless stated otherwise. Uncertainties without unit are relative uncertainties referring to the measurement value unless stated otherwise.

Annex to the accreditation certificate D-K-17601-01-00

On-site Calibration

Calibration and Measurement Capabilities (CMC)

Measurement quantity / Calibration item	Range	Measurement conditions / procedure	Expanded uncertainty of measurement ¹⁾	Remarks
DC voltage	-1 kV to -100 kV		0.7 %	
	1 kV to 100 kV		0.7 %	
AC voltage	1 kV to 100 kV	50 Hz	0.4 %	
	50 kV to 300 kV	50 Hz	0.8 %	
Burst voltage (LI) Amplitude	-10 kV to -500 kV		0.8 %	LI = Lightning Impulse T_1 = Front slope time T_2 = Back slope-half lifetime
	10 kV to 500 kV		0.8 %	
Time parameter T_1	0.8 μ s to 1.6 μ s		2.2 %	
	T_2 40 μ s to 60 μ s		2.2 %	

Abbreviations used:

CMC Calibration and measurement capabilities
DIN Deutsches Institut für Normung e.V.

¹⁾ The expanded uncertainties according to EA-4/02 M:2013 are part of CMC and are the best measurement uncertainties within accreditation. They have a coverage probability of approximately 95 % and have a coverage factor of $k = 2$ unless stated otherwise. Uncertainties without unit are relative uncertainties referring to the measurement value unless stated otherwise.