

Deutsche Akkreditierungsstelle GmbH German Accreditation Body

Annex to the Accreditation Certificate D-K-18351-01-00 according to DIN EN ISO/IEC 17025:2005

Valid from: 2018-10-23

Date of issue: 2018-12-04

Holder of certificate:

ZwickRoell GmbH & Co. KG
August-Nagel-Straße 11, 89079 Ulm

with the further locations:

**Parc Empresarial Trade Center, Avda Corts Catalanes 5-7 planta 2a Local 1,
E-08173 Santa Cugat del Valles (Barcelona), Spain**

18 Boon Lay Way, #06-105/106, TradeHub 21, Singapore 609966

Corso Perrone 39 h rosso, I-16152 Genova, Italy

Head: Dipl.-Ing. (FH) Thomas Gaube
Deputy: Oliver Glökler
Dipl.-Ing. Stephan Baumann

Accredited as calibration laboratory since: 1994-09-27

Calibrations in the fields:

Mechanical quantities

Material testing machines (MTM)

- Force (MTM) ^{a)}
- Extension (MTM) ^{a)}
- Mechanical work (MTM) ^{a)}
- Hardness (MTM) ^{a)}
- Torque (MTM) ^{a)}
- Angle of rotation (MTM) ^{a)}
- Velocity (MTM) ^{a)}

Thermodynamic quantities

Temperature quantities

- Climatic chambers (temperature) ^{a)}
- Thermocouples ^{a)}

^{a)} only on site calibration

Abbreviations used: see last page

Within the measurands / calibration items marked with ^{a)}, the calibration laboratory is permitted, without being required to inform and obtain prior approval from DAkkS, to use calibration standards or equivalent calibration procedures listed here with different issue dates. The calibration laboratory maintains a current list of all calibration standards / equivalent calibration procedures within the flexible scope of accreditation.

Annex to the accreditation certificate D-K-18351-01-00

August-Nagel-Straße 11, 89079 Ulm

On-site calibration

Measured quantity / Calibration item	Range	Measurement conditions / procedure	Best measurement capability ¹⁾	Remarks
Force (MTM) ^{*)} Force measuring devices of materials testing machines according to DIN 51220	1 N to 600 kN	DIN EN ISO 7500-1:2018 with supplementary sheet 1 to 3:1999 supplementary sheet 4:2013	0.12 %	Force transducer Class 0.5 tensile / compression
	200 N to 3000 kN		0.12 %	Force transducer Class 0.5 tensile
	200 N to 5000 kN	DIN EN ISO 7500-2:2007 DIN EN ISO 6506-2:2015 DIN EN ISO 6507-2:2013	0.12 %	Force transducer Class 0.5 compression
	0.02 N to 200 N	DIN EN ISO 6508-2:2015 DIN EN ISO 4545-2:2015 DIN EN ISO 2039-1:2003 DIN EN ISO 14577-2:2015 ASTM E10:2017 ASTM E92:2017 ASTM E384:2017 ASTM E18:2017 ASTM F36:2015 ASTM E4:2016	0.10 %	Known masses tensile / tensile and compression
Extension (MTM) ^{*)} Extension measuring devices of materials testing machines according to DIN 51220	0 mm to 60 mm	DIN EN ISO 9513:2013 DIN EN ISO 6508-2:2015 DIN EN ISO 2039-1:2003 DIN EN ISO 14577-2:2015 DIN EN ISO 527:2012 ASTM F36:2015 ASTM E83:2016 ASTM E2309:2016	1.5 · 10 ⁻³ · l; but not <0.5 µm	Measuring principle: incremental probe (CP60/CT6002)
	0 mm to 12 mm		2 · 10 ⁻³ · l; but not <2 µm	Measuring principle: incremental probe (MT 12)
	0 mm to 12 mm		1.5 · 10 ⁻³ · l; but not <0.5 µm	Measuring principle: incremental probe (MT 1201)
	0 mm to 205 mm		2 · 10 ⁻³ · l; but not < 5 µm	Measuring principle: probe on basis of magnet technology
	0 mm to 1500 mm		2 · 10 ⁻³ · l; but not <4 µm	Measuring principle: Rotary encoder with incremental divide
	0.1 mm to 100 mm		1.5 · 10 ⁻³ · l; but not <0.5 µm	Gauge blocks class 1
Optical indentation measuring devices of Hardness Testers	0 mm to 6 mm	DIN EN ISO 6506-2:2015 DIN EN ISO 6507-2:2013 DIN EN ISO 4545-2:2015 ASTM E10:2017 ASTM E92:2017 ASTM E384:2017	1.5 · 10 ⁻³ · l; but not <0.5 µm	Measuring principle: Object micrometer in incident light
Depth measuring device of Hardness Testers	0 mm to 0.8 mm	DIN EN ISO 6508-2:2015 ASTM E18:2017	1.5 · 10 ⁻³ · l; but not <0.5 µm	Measuring principle: incremental probe

¹⁾ The best measurement capabilities are stated according to DAkkS-DKD-3 (EA-4/02). These are expanded uncertainties of measurement with a coverage probability of 95 % and have a coverage factor of $k = 2$ unless stated otherwise. Uncertainties without unit are relative uncertainties referring to the measurement value unless stated otherwise.

Annex to the accreditation certificate D-K-18351-01-00

August-Nagel-Straße 11, 89079 Ulm

On-site calibration

Measured quantity / Calibration item	Range	Measurement conditions / procedure	Best measurement capability ¹⁾	Remarks
Mechanical Work (MTM) ¹⁾ Pendulum Impact Testers and Impact Testing Devices	0.2 J to 750 J	DIN 51222:2017 DIN 51230:1977 DIN 53512:2000 DIN EN-ISO 148-2:2017 DIN EN ISO 13802:2016 ASTM E23:2016b	Force: 0.12 % Pendulum length: 0.17 mm Angle: 0.03° Time: 0.02 s	The measurement uncertainty is calculated for: 1. Position of centre of percussion, 2. Potential energy, 3. Error of the indicated energy.
Hardness (MTM) ¹⁾ Hardness Testers according to Brinell-, Vickers-, Rockwell-, Knoop- and Martens test procedure	100 HB to 550 HB	DIN EN ISO 6506-2:2015 DIN EN ISO 6507-2:2013 DIN EN ISO 6508-2:2015 DIN EN ISO 4545-2:2015 DIN EN ISO 2039-1:2003 DIN EN ISO 14577-2:2015 ASTM E10:2017 ASTM E92:2017 ASTM E384:2017 ASTM E18:2017 ASTM F36:2015	2 % HB	The values indicated for the measurement uncertainty are valid for the indirect verification with hardness comparison plates. The measurement uncertainty of the individual parameters of the direct verification is indicated separately. U_{CRM} = calibration uncertainty of the hardness comparison plate
	30 HV to 950 HV (Hardness scales HV5 to HV100) (Hardness scales HV 0.01 to HV3)		1 % HV, but not < $1,5 \cdot U_{CRM}$ 2 % HV, but not < $1,5 \cdot U_{CRM}$	
	100 HK to 950 HK (Hardness scales HK 0.01 to HK 2)		2 % HK, but not < $1,5 \cdot U_{CRM}$	
	20 HRA to 65 HRA		1.0 HRA	
	66 HRA to 95 HRA		0.5 HRA	
	10 HRB to 55 HRB		1.5 HRB	
	56 HRB to 100 HRB		1.0 HRB	
	20 HRC to 55 HRC		1.0 HRC	
	56 HRC to 70 HRC		0.5 HRC	
	40 HRD to 69 HRD		1.5 HRD	
	70 HRD to 77 HRD		1.0 HRD	
	60 HRF to 100 HRF		1.0 HRF	
	20 HRN to 60 HRN		1.0 HRN	
	61 HRN to 91 HRN		0.5 HRN	
12 HRT to 93 HRT	2.0 HRT			
Torque (MTM) Torque measuring devices of materials testing machines according to DIN 51220	0.2 N·m to 2000 N·m	QI-D-005: 2018	0.30 %	With torque transducers (clockwise and counterclockwise torque)
	0.02 N·m to 20 N·m		0.30 %	Known masses tensile in combination with lever arm

¹⁾ The best measurement capabilities are stated according to DAKKS-DKD-3 (EA-4/02). These are expanded uncertainties of measurement with a coverage probability of 95 % and have a coverage factor of $k = 2$ unless stated otherwise. Uncertainties without unit are relative uncertainties referring to the measurement value unless stated otherwise.

Annex to the accreditation certificate D-K-18351-01-00
August-Nagel-Straße 11, 89079 Ulm
On-site calibration

Measured quantity / Calibration item	Range	Measurement conditions / procedure	Best measurement capability ¹⁾	Remarks
Velocity (MTM) ¹⁾ Traverse speed of materials testing machines according to DIN 51220	0.1 mm/min to 500 mm/min	ASTM E2658:2015	1.0 %	Measuring principle: Start/Stop-Method of distance and time
Angle of rotation (MTM) Measuring devices for angle of rotation on materials testing machines according to DIN 51220	1° to 360°	QI-D-006: 2018	$3 \cdot 10^{-3} \cdot W$	Measuring principle: incremental <i>W</i> : measured angle
Temperature ¹⁾ Climate chambers Climate chamber with air circulation in empty or defined loaded usable space	-80 °C to -40 °C	DAkks-DKD-R 5-7: 2010 Method C Measurement in air	0.2 K	Comparison with standard thermometers
	> -40 °C to 0 °C		0.15 K	
	> 0 °C to 100 °C		0.10 K	
	> 100 °C to 150 °C		0.15 K	
	> 150 °C to 200 °C		0.25 K	
	> 200 °C to 250 °C		0.35 K	
Climate chamber with air circulation in empty or defined loaded usable space	-80 °C to -40 °C	DAkks-DKD-R 5-7: 2010 Method A and B Measurement in air	0.5 K	Comparison with standard thermometers
	> -40 °C to 0 °C		0.4 K	
	> 0 °C to 100 °C		0.2 K	
	> 100 °C to 150 °C		0.4 K	
	> 150 °C to 200 °C		0.6 K	
	> 200 °C to 250 °C		1.7 K	
Climate chamber without air circulation in empty or defined loaded usable space	-80 °C to -40 °C	DAkks-DKD-R 5-7: 2010 Method C Measurement in air	0.5 K	Comparison with standard thermometers
	> -40 °C to 0 °C		0.4 K	
	> 0 °C to 100 °C		0.3 K	
	> 100 °C to 150 °C		0.4 K	
	> 150 °C to 200 °C		0.5 K	
	> 200 °C to 250 °C		0.8 K	

¹⁾ The best measurement capabilities are stated according to DAkkS-DKD-3 (EA-4/02). These are expanded uncertainties of measurement with a coverage probability of 95 % and have a coverage factor of $k = 2$ unless stated otherwise. Uncertainties without unit are relative uncertainties referring to the measurement value unless stated otherwise.

Annex to the accreditation certificate D-K-18351-01-00

August-Nagel-Straße 11, 89079 Ulm

On-site calibration

Measured quantity / Calibration item	Range	Measurement conditions / procedure	Best measurement capability ¹⁾	Remarks
Temperature ¹⁾ Climate chambers Climate chamber without air circulation in empty or defined loaded usable space	-80 °C to -40 °C	DAkKS-DKD-R 5-7: 2010 Method A and B Measurement in air	3.0 K	Comparison with standard thermometers
	> -40 °C to 0 °C		2.0 K	
	> 0 °C to 100 °C		2.2 K	
	> 100 °C to 150 °C		3.0 K	
	> 150 °C to 200 °C		3.5 K	
	> 200 °C to 250 °C		5.0 K	
Thermocouples with indicator device	150 °C to 300 °C	DAkKS-DKD-R 5-3: 2010 In block calibrator Pegasus	2.8 K	Comparison with standard thermometers
	> 300 °C to 600 °C		3.5 K	
	> 600 °C to 900 °C		4.3 K	
	> 900 °C to 1200 °C		5.5 K	

¹⁾ The best measurement capabilities are stated according to DAkkS-DKD-3 (EA-4/02). These are expanded uncertainties of measurement with a coverage probability of 95 % and have a coverage factor of $k = 2$ unless stated otherwise. Uncertainties without unit are relative uncertainties referring to the measurement value unless stated otherwise.

Annex to the accreditation certificate D-K-18351-01-00

Parc Empresarial Trade Center, Avda Corts Catalanes 5-7 planta 2a Local 1, E-08173 Santa Cugat del Valles (Barcelona), Spain

On-site calibration

Measured quantity / Calibration item	Range	Measurement conditions / procedure	Best measurement capability ¹⁾	Remarks
Force (MTM) ^{*)} Force measuring devices of materials testing machines according to DIN 51220	1 N to 600 kN	DIN EN ISO 7500-1:2018 with supplementary sheet 1 to 3:1999	0.12 %	Force transducer Class 0.5 tensile / compression
	200 N to 3000 kN	supplementary sheet 4:2013	0.12 %	Force transducer Class 0.5 tensile
	200 N to 5000 kN	DIN EN ISO 7500-2:2007 DIN EN ISO 6506-2:2015 DIN EN ISO 6507-2:2013	0.12 %	Force transducer Class 0.5 compression
	0.02 N to 200 N	DIN EN ISO 6508-2:2015 DIN EN ISO 4545-2:2015 DIN EN ISO 2039-1:2003 ASTM E10:2017 ASTM E92:2017 ASTM E384:2017 ASTM E18:2017 ASTM F36:2015 ASTM E4:2016	0.10 %	Known masses tensile / compression
Extension (MTM) ^{*)} Extension measuring devices of materials testing machines according to DIN 51220	0 mm to 60 mm	DIN EN ISO 9513:2013 DIN EN ISO 6508-2:2015 DIN EN ISO 2039-1:2003 DIN EN ISO 527-1:2012	$1.5 \cdot 10^{-3} \cdot l$; but not $<0.5 \mu\text{m}$	Measuring principle: incremental probe (CP60/CT6002)
	0 mm to 12 mm	ASTM F36:2015 ASTM E83:2016 ASTM E2309:2016	$2 \cdot 10^{-3} \cdot l$; but not $<2 \mu\text{m}$	Measuring principle: incremental probe (MT 12)
	0 mm to 12 mm		$1.5 \cdot 10^{-3} \cdot l$; but not $<0.5 \mu\text{m}$	Measuring principle: incremental probe (MT 1201)
	0 mm to 205 mm		$2 \cdot 10^{-3} \cdot l$; but not $<5 \mu\text{m}$	Measuring principle: probe on basis of magnet technology
	0 mm to 1500 mm		$2 \cdot 10^{-3} \cdot l$; but not $<4 \mu\text{m}$	Measuring principle: Rotary encoder with incremental divide
	0.1 mm to 100 mm		$1.5 \cdot 10^{-3} \cdot l$; but not $<0.5 \mu\text{m}$	Gauge blocks class 1
Optical indentation measuring devices of Hardness Testers	0 mm to 6 mm	DIN EN ISO 6506-2:2015 DIN EN ISO 6507-2:2013 DIN EN ISO 4545-2:2015 ASTM E10:2017 ASTM E92:2017 ASTM E384:2017	$1.5 \cdot 10^{-3} \cdot l$; but not $<0.5 \mu\text{m}$	Measuring principle: Object micrometer in incident light
Depth measuring device of Hardness Testers	0 mm to 0.8 mm	DIN EN ISO 6508-2:2015 ASTM E18:2017	$1.5 \cdot 10^{-3} \cdot l$; but not $<0.5 \mu\text{m}$	Measuring principle: incremental probe

¹⁾ The best measurement capabilities are stated according to DAkks-DKD-3 (EA-4/02). These are expanded uncertainties of measurement with a coverage probability of 95 % and have a coverage factor of $k = 2$ unless stated otherwise. Uncertainties without unit are relative uncertainties referring to the measurement value unless stated otherwise.

Parc Empresarial Trade Center, Avda Corts Catalanes 5-7 planta 2a Local 1, E-08173 Santa Cugat del Valles (Barcelona), Spain
On-site calibration

Measured quantity / Calibration item	Range	Measurement conditions / procedure	Best measurement capability ¹⁾	Remarks
Mechanical Work (MTM) ¹⁾ Pendulum Impact Testers and Impact Testing Devices	0.2 J to 750 J	DIN 51222:2017 DIN 51230:1977 DIN 53512:2000 DIN EN-ISO 148-2:2017 DIN EN ISO 13802:2016 ASTM E23:2016b	Force: 0.12 % Pendulum length: 0.17 mm Angle: 0.03° Time: 0.02 s	The measurement uncertainty is calculated for: 1. Position of centre of percussion, 2. Potential energy, 3. Error of the indicated energy.
Hardness (MTM) ¹⁾ Hardness Testers according to Brinell-, Vickers-, Knoop-, and Rockwell test procedure	100 HB to 550 HB	DIN EN ISO 6506-2:2015 DIN EN ISO 6507-2:2013 DIN EN ISO 6508-2:2015 DIN EN ISO 4545-2:2015 DIN EN ISO 2039-1:2003 ASTM E10:2017 ASTM E92:2017 ASTM E384:2017 ASTM E18:2017 ASTM F36:2015	2 % HB	The values indicated for the measurement uncertainty are valid for the indirect verification with hardness comparison plates. The measurement uncertainty of the individual parameters of the direct verification is indicated separately. U_{CRM} = calibration uncertainty of the hardness comparison plate
30 HV to 950 HV (Hardness scales HV5 to HV100) (Hardness scales HV 0.01 to HV3)	1 % HV, but not < 1,5 · U_{CRM} 2 % HV, but not < 1.5 · U_{CRM}			
100 HK to 950 HK (Hardness scales HK 0.01 to HK 2)	2 % HK, but not < 1.5 · U_{CRM}			
20 HRA to 65 HRA	1.0 HRA			
66 HRA to 95 HRA	0.5 HRA			
10 HRB to 55 HRB	1.5 HRB			
56 HRB to 100 HRB	1.0 HRB			
20 HRC to 55 HRC	1.0 HRC			
56 HRC to 70 HRC	0.5 HRC			
40 HRD to 69 HRD	1.5 HRD			
70 HRD to 77 HRD	1.0 HRD			
60 HRF to 100 HRF	1.0 HRF			
20 HRN to 60 HRN	1.0 HRN			
61 HRN to 91 HRN	0.5 HRN			
12 HRT to 93 HRT	2.0 HRT			

¹⁾ The best measurement capabilities are stated according to DAkks-DKD-3 (EA-4/02). These are expanded uncertainties of measurement with a coverage probability of 95 % and have a coverage factor of $k = 2$ unless stated otherwise. Uncertainties without unit are relative uncertainties referring to the measurement value unless stated otherwise.

Annex to the accreditation certificate D-K-18351-01-00

18 Boon Lay Way, #06-105/106, TradeHub 21, Singapore 609966

On-site calibration

Measured quantity / Calibration item	Range	Measurement conditions / procedure	Best measurement capability ¹⁾	Remarks
Force (MTM) ¹⁾ Force measuring devices of materials testing machines according to DIN 51220	1 N to 600 kN	DIN EN ISO 7500-1:2018 with supplementary sheet 1 to 3:1999 supplementary sheet 4:2013 DIN EN ISO 7500-2:2007 ASTM E4:2016	0.12 %	Force transducer Class 0.5 tensile / compression
	200 N to 3000 kN		0.12 %	Force transducer Class 0.5 tensile
	200 N to 5000 kN		0.12 %	Force transducer Class 0.5 compression
	0.02 N to 200 N		0.10 %	Known masses tensile / compression
Extension (MTM) ¹⁾ Extension measuring devices of materials testing machines according to DIN 51220	0 mm to 60 mm	DIN EN ISO 9513:2013 DIN EN ISO 527-1:2012 ASTM E83:2016 ASTM E2309:2016	$1.5 \cdot 10^{-3} \cdot l$; but not $<0.5 \mu\text{m}$	Measuring principle: incremental probe (CP60/CT6002)
	0 mm to 12 mm		$2 \cdot 10^{-3} \cdot l$; but not $<2 \mu\text{m}$	Measuring principle: incremental probe (MT 12)
	0 mm to 12 mm		$1.5 \cdot 10^{-3} \cdot l$; but not $<0.5 \mu\text{m}$	Measuring principle: incremental probe (MT 1201)
	0 mm to 205 mm		$2 \cdot 10^{-3} \cdot l$; but not $<5 \mu\text{m}$	Measuring principle: probe on basis of magnet technology
	0 mm to 1500 mm		$2 \cdot 10^{-3} \cdot l$; but not $<4 \mu\text{m}$	Measuring principle: Rotary encoder with incremental divide
	0.1 mm to 100 mm		$1.5 \cdot 10^{-3} \cdot l$; but not $<0.5 \mu\text{m}$	Gauge blocks class 1
Mechanical Work (MTM) ¹⁾ Pendulum Impact Testers and Impact Testing Devices	0.2 J to 750 J	DIN 51222:2017 DIN 51230:1977 DIN 53512:2000 DIN EN-ISO 148-2:2017 DIN EN ISO 13802:2016 ASTM E23:2016b	Force: 0.12 % Pendulum length: 0.17 mm Angle: 0.03° Time: 0.02 s	The measurement uncertainty is calculated for: 1. Position of centre of percussion, 2. Potential energy, 3. Error of the indicated energy.

¹⁾ The best measurement capabilities are stated according to DAKKS-DKD-3 (EA-4/02). These are expanded uncertainties of measurement with a coverage probability of 95 % and have a coverage factor of $k = 2$ unless stated otherwise. Uncertainties without unit are relative uncertainties referring to the measurement value unless stated otherwise.

Annex to the accreditation certificate D-K-18351-01-00

Corso Perrone 39 h rosso, I-16152 Genova, Italy

On-site calibration

Measured quantity / Calibration item	Range	Measurement conditions / procedure	Best measurement capability ¹⁾	Remarks
Force (MTM) ¹⁾ Force measuring devices of materials testing machines according to DIN 51220	1 N to 600 kN	DIN EN ISO 7500-1:2018 with supplementary sheet 1 to 3:1999 supplementary sheet 4:2013	0.12 %	Force transducer Class 0.5 tensile / compression
	200 N to 3000 kN	DIN EN ISO 7500-2:2007 ASTM E4:2016	0.12 %	Force transducer Class 0.5 tensile
	200 N to 5000 kN	DIN EN ISO 7500-2:2007 ASTM E4:2016	0.12 %	Force transducer Class 0.5 compression
	0.02 N to 200 N	DIN EN ISO 6506-2:2015 DIN EN ISO 6507-2:2013 DIN EN ISO 6508-2:2015 DIN EN ISO 4545-2:2015 DIN EN ISO 2039-1:2003 ASTM E10:2017 ASTM E92:2017 ASTM E384:2017 ASTM E18:2017 ASTM F36:2015	0.10 %	Known masses tensile / compression
Extension (MTM) ¹⁾ Extension measuring devices of materials testing machines according to DIN 51220	0 mm to 60 mm	DIN EN ISO 9513:2013 DIN EN ISO 6508-2:2015 DIN EN ISO 2039-1:2003	$1.5 \cdot 10^{-3} \cdot l$; but not $<0.5 \mu\text{m}$	Measuring principle: incremental probe (CP60/CT6002)
	0 mm to 12 mm	DIN EN ISO 527-1:2012 ASTM E83:2016 ASTM E2309:2016 ASTM F36:2015	$2 \cdot 10^{-3} \cdot l$; but not $<2 \mu\text{m}$	Measuring principle: incremental probe (MT 12)
	0 mm to 12 mm		$1.5 \cdot 10^{-3} \cdot l$; but not $<0.5 \mu\text{m}$	Measuring principle: incremental probe (MT 1201)
	0 mm to 205 mm		$2 \cdot 10^{-3} \cdot l$; but not $<5 \mu\text{m}$	Measuring principle: probe on basis of magnet technology
	0 mm to 1500 mm		$2 \cdot 10^{-3} \cdot l$; but not $<4 \mu\text{m}$	Measuring principle: Rotary encoder with incremental divide
	0.1 mm to 100 mm		$1.5 \cdot 10^{-3} \cdot l$; but not $<0.5 \mu\text{m}$	Gauge blocks class 1
Optical indentation measuring devices of Hardness Testers	0 mm to 6 mm	DIN EN ISO 6506-2:2015 DIN EN ISO 6507-2:2013 DIN EN ISO 4545-2:2015 ASTM E10:2017 ASTM E92:2017 ASTM E384:2017	$1.5 \cdot 10^{-3} \cdot l$; but not $<0.5 \mu\text{m}$	Measuring principle: Object micrometer in incident light
Depth measuring device of Hardness Testers	0 mm to 0.8 mm	DIN EN ISO 6508-2:2015 ASTM E18:2017	$1.5 \cdot 10^{-3} \cdot l$; but not $<0.5 \mu\text{m}$	Measuring principle: incremental probe

¹⁾ The best measurement capabilities are stated according to DAkks-DKD-3 (EA-4/02). These are expanded uncertainties of measurement with a coverage probability of 95 % and have a coverage factor of $k = 2$ unless stated otherwise. Uncertainties without unit are relative uncertainties referring to the measurement value unless stated otherwise.

Annex to the accreditation certificate D-K-18351-01-00
Corso Perrone 39 h rosso, I-16152 Genova, Italy
On-site calibration

Measured quantity / Calibration item	Range	Measurement conditions / procedure	Best measurement capability ¹⁾	Remarks
Mechanical Work (MTM) ¹⁾ Pendulum Impact Testers and Impact Testing Devices	0.2 J to 750 J	DIN 51222:2017 DIN 51230:1977 DIN 53512:2000 DIN EN-ISO 148-2:2017 DIN EN ISO 13802: 2016 ASTM E23:2016b	Force: 0.12 % Pendulum length: 0.17 mm Angle: 0.03° Time: 0.02 s	The measurement uncertainty is calculated for: 1. Position of centre of percussion, 2. Potential energy, 3. Error of the indicated energy.
Hardness (MTM) ¹⁾ Hardness Testers according to Brinell-, Vickers-, Knoop-, and Rockwell test procedure	100 HB to 550 HB	DIN EN ISO 6506-2:2015 DIN EN ISO 6507-2:2013 DIN EN ISO 6508-2:2015 DIN EN ISO 4545-2:2015 DIN EN ISO 2039-1:2003 ASTM E10:2017 ASTM E92:2017 ASTM E384:2017 ASTM E18:2017 ASTM F36:2015	2 % HB	The values indicated for the measurement uncertainty are valid for the indirect calibration with hardness comparison plates. The measurement uncertainty of the individual parameters of the direct calibration is indicated separately. (U_{CRM} = calibration uncertainty of the hardness comparison plate)
	30 HV to 950 HV (Hardness scales HV5 to HV100) (Hardness scales HV0.01 to HV3)		1 % HV, but not $< 1.5 \cdot U_{CRM}$ 2 % HV, but not $< 1.5 \cdot U_{CRM}$	
	100 HK to 950 HK (Hardness scale HK 0.01 to HK 2)		2 % HK, but not $< 1.5 \cdot U_{CRM}$	
	20 HRA to 65 HRA		1.0 HRA	
	66 HRA to 95 HRA		0.5 HRA	
	10 HRB to 55 HRB		1.5 HRB	
	56 HRB to 100 HRB		1.0 HRB	
	20 HRC to 55 HRC		1.0 HRC	
	56 HRC to 70 HRC		0.5 HRC	
	40 HRD to 69 HRD		1.5 HRD	
	70 HRD to 77 HRD		1.0 HRD	
	60 HRF to 100 HRF		1.0 HRF	
	20 HRN to 60 HRN		1.0 HRN	
61 HRN to 91 HRN	0.5 HRN			
12 HRT to 93 HRT	2.0 HRT			

¹⁾ The best measurement capabilities are stated according to DAkks-DKD-3 (EA-4/02). These are expanded uncertainties of measurement with a coverage probability of 95 % and have a coverage factor of $k=2$ unless stated otherwise. Uncertainties without unit are relative uncertainties referring to the measurement value unless stated otherwise.

Annex to the accreditation certificate D-K-18351-01-00

Abbreviations used:

ASTM	ASTM American Standard for Testing and Materials
DIN	Deutsches Institut für Normung e.V. (German Institut for Standardization)
EN	European Standard
ISO	International Organisation for Standardization
QI	Quality Instruction (In house calibration procedure of calibration laboratory)

¹⁾ The best measurement capabilities are stated according to DAkkS-DKD-3 (EA-4/02). These are expanded uncertainties of measurement with a coverage probability of 95 % and have a coverage factor of $k = 2$ unless stated otherwise. Uncertainties without unit are relative uncertainties referring to the measurement value unless stated otherwise.